

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: WOOD06
Lab Code: CHEM Case No.: Q1669 SAS No.: Q1669 SDG NO.: Q1669
Lab File ID: VL042200.D BFB Injection Date: 03/31/2025
Instrument ID: MSVOA_L BFB Injection Time: 08:09
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	25.4
75	30.0 - 66.0% of mass 95	61.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.3 (0.5) 1
174	50.0 - 120.0% of mass 95	59.4
175	4.0 - 9.0% of mass 174	4.5 (7.6) 1
176	93.0 - 101.0% of mass 174	57.9 (97.6) 1
177	5.0 - 9.0% of mass 176	3.8 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL042201.D	03/31/2025	08:51
VL0331ABL01	VL0331ABL01	VL042202.D	03/31/2025	09:34
VL0331ABS01	VL0331ABS01	VL042203.D	03/31/2025	10:17
SSSG-DUP-1-3262025	Q1669-06	VL042204.D	03/31/2025	11:06
SSSG-DUP-1-3262025DL	Q1669-06DL	VL042205.D	03/31/2025	11:52
SSSG-DUP-1-3262025DUP	Q1669-06DUP	VL042206.D	03/31/2025	12:26