

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: APTIM **SDG No.:** Q1673
Contract: SHAW18 **Lab Code:** CHEM **Case No.:** Q1673 **SAS No.:** Q1673
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Selenium	1030	1000	103	95 - 105	P	04/01/2025	10:57	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: APTIM **SDG No.:** Q1673
Contract: SHAW18 **Lab Code:** CHEM **Case No.:** Q1673 **SAS No.:** Q1673
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Selenium	21.8	20.0	109	80 - 120	P	04/01/2025	11:15	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: APTIM **SDG No.:** Q1673
Contract: SHAW18 **Lab Code:** CHEM **Case No.:** Q1673 **SAS No.:** Q1673
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Selenium	5090	5000	102	90 - 110	P	04/01/2025	11:50	LB135255
CCV02	Selenium	5180	5000	104	90 - 110	P	04/01/2025	12:43	LB135255
CCV03	Selenium	5390	5000	108	90 - 110	P	04/01/2025	13:35	LB135255
CCV04	Selenium	5300	5000	106	90 - 110	P	04/01/2025	14:31	LB135255
CCV05	Selenium	5210	5000	104	90 - 110	P	04/01/2025	15:23	LB135255
CCV06	Selenium	5180	5000	104	90 - 110	P	04/01/2025	15:52	LB135255
CCV07	Selenium	5160	5000	103	90 - 110	P	04/01/2025	16:56	LB135255
CCV08	Selenium	5280	5000	106	90 - 110	P	04/01/2025	17:41	LB135255