



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Metals**

**- 2b -**

**CRDL STANDARD FOR AA & ICP**

**Client:** APTIM **SDG No.:** Q1673  
**Contract:** SHAW18 **Lab Code:** CHEM **Case No.:** Q1673 **SAS No.:** Q1673  
**Initial Calibration Source:** \_\_\_\_\_  
**Continuing Calibration Source:** \_\_\_\_\_

| Sample ID | Analyte  | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|--------------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CRI01     | Selenium | 19.7           | 20.0               | 98            | 40 - 160                  | P | 04/01/2025       | 11:28            | LB135255      |