



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	APTIM				SDG No.:	Q1673			
Contract:	SHAW18	Lab Code:	CHEM		Case No.:	Q1673	SAS No.:	Q1673	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	11:19	LB135255

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: APTIM SDG No.: Q1673
Contract: SHAW18 Lab Code: CHEM Case No.: Q1673 SAS No.: Q1673

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	11:54	LB135255
CCB02	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	12:48	LB135255
CCB03	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	13:39	LB135255
CCB04	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	14:36	LB135255
CCB05	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	15:28	LB135255
CCB06	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	15:56	LB135255
CCB07	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	17:00	LB135255
CCB08	Selenium	20.0	+/-20.0	U	20.0	P	04/01/2025	17:50	LB135255