

**Metals**

- 4 -

**INTERFERENCE CHECK SAMPLE**

**Client:** APTIM **SDG No.:** Q1673  
**Contract:** SHAW18 **Lab Code:** CHEM **Case No.:** Q1673 **SAS No.:** Q1673  
**ICS Source:** EPA **Instrument ID:** P4

| Sample ID | Analyte  | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| ICSA01    | Selenium | -13.6          |                    |               | -20                    | 20                      | 04/01/2025       | 11:33            | LB135255      |
| ICSAB01   | Selenium | 39.8           | 46.0               | 86            | 26                     | 66                      | 04/01/2025       | 11:37            | LB135255      |
| ICSA      | Selenium | 22.0           |                    |               | -20                    | 20                      | 04/01/2025       | 11:41            | LB135255      |
| ICSAB     | Selenium | 60.6           | 46.0               | 132           | 26                     | 66                      | 04/01/2025       | 11:46            | LB135255      |