

Continuing Calibration Report for SequenceID : FE033125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053070.D

Aliphatic C9-C12	10101647.000	60.000	3.141	6.784	168360.783	156032.227	-7.901
Aliphatic C12-C16	6705782.000	40.000	6.785	10.235	167644.550	153631.377	-9.121
Aliphatic C16-C21	9524747.000	60.000	10.236	13.612	158745.783	144337.201	-9.983
Aliphatic C21-C28	11830233.000	80.000	13.613	17.286	147877.913	134347.450	-10.071
Aliphatic C28-C40	15356062.000	120.000	17.287	22.206	127967.183	124256.607	-2.986
Aliphatic EPH	53518471.000	360.000	3.141	22.206	148662.419	138405.582	-7.411