

Continuing Calibration Report for SequenceID : FE033125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053086.D

Aliphatic C9-C12	8865929.000	60.000	3.141	6.784	147765.483	156032.227	5.298
Aliphatic C12-C16	6037117.000	40.000	6.785	10.235	150927.925	153631.377	1.760
Aliphatic C16-C21	8611980.000	60.000	10.236	13.612	143533.000	144337.201	0.557
Aliphatic C21-C28	11243627.000	80.000	13.613	17.286	140545.338	134347.450	-4.613
Aliphatic C28-C40	16692620.000	120.000	17.287	22.206	139105.167	124256.607	-11.950
Aliphatic EPH	51451273.000	360.000	3.141	22.206	142920.203	138405.582	-3.262