

Continuing Calibration Report for SequenceID : FE040125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053089.D

Aliphatic C9-C12	9641759.000	60.000	3.141	6.782	160695.983	156032.227	-2.989
Aliphatic C12-C16	6370470.000	40.000	6.783	10.232	159261.750	153631.377	-3.665
Aliphatic C16-C21	9039424.000	60.000	10.233	13.609	150657.067	144337.201	-4.379
Aliphatic C21-C28	11323535.000	80.000	13.610	17.278	141544.188	134347.450	-5.357
Aliphatic C28-C40	15436560.000	120.000	17.279	22.189	128638.000	124256.607	-3.526
Aliphatic EPH	51811748.000	360.000	3.141	22.189	143921.522	138405.582	-3.985