

Continuing Calibration Report for SequenceID : FE040125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE053093.D

Aliphatic C9-C12	9633869.000	60.000	3.141	6.782	160564.483	156032.227	-2.905
Aliphatic C12-C16	6368244.000	40.000	6.783	10.232	159206.100	153631.377	-3.629
Aliphatic C16-C21	9013032.000	60.000	10.233	13.609	150217.200	144337.201	-4.074
Aliphatic C21-C28	11327743.000	80.000	13.610	17.278	141596.788	134347.450	-5.396
Aliphatic C28-C40	15471341.000	120.000	17.279	22.189	128927.842	124256.607	-3.759
Aliphatic EPH	51814229.000	360.000	3.141	22.189	143928.414	138405.582	-3.990