

Continuing Calibration Report for SequenceID : FC033125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068444.D

Aliphatic C9-C12	9465089.000	60.000	3.098	6.359	157751.483	162319.116	2.814
Aliphatic C12-C16	6556350.000	40.000	6.360	9.743	163908.750	168194.168	2.548
Aliphatic C16-C21	10024155.000	60.000	9.744	13.097	167069.250	171510.591	2.590
Aliphatic C21-C28	12608586.000	80.000	13.098	16.748	157607.325	168869.942	6.669
Aliphatic C28-C40	14864058.000	120.000	16.749	21.546	123867.150	136171.748	9.036
Aliphatic EPH	53518238.000	360.000	3.098	21.546	148661.772	157243.762	5.458