

Continuing Calibration Report for SequenceID : FC033125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068461.D

Aliphatic C9-C12	10317792.000	60.000	3.098	6.359	171963.200	162319.116	-5.941
Aliphatic C12-C16	7151887.000	40.000	6.360	9.743	178797.175	168194.168	-6.304
Aliphatic C16-C21	10644632.000	60.000	9.744	13.097	177410.533	171510.591	-3.440
Aliphatic C21-C28	13003247.000	80.000	13.098	16.748	162540.588	168869.942	3.748
Aliphatic C28-C40	15298197.000	120.000	16.749	21.546	127484.975	136171.748	6.379
Aliphatic EPH	56415755.000	360.000	3.098	21.546	156710.431	157243.762	0.339