

Continuing Calibration Report for SequenceID : FC040125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068464.D

Aliphatic C9-C12	9793802.000	60.000	3.097	6.359	163230.033	162319.116	-0.561
Aliphatic C12-C16	6722165.000	40.000	6.360	9.742	168054.125	168194.168	0.083
Aliphatic C16-C21	9852986.000	60.000	9.743	13.096	164216.433	171510.591	4.253
Aliphatic C21-C28	12280464.000	80.000	13.097	16.749	153505.800	168869.942	9.098
Aliphatic C28-C40	14868463.000	120.000	16.750	21.544	123903.858	136171.748	9.009
Aliphatic EPH	53517880.000	360.000	3.097	21.544	148660.778	157243.762	5.458