

Continuing Calibration Report for SequenceID : FC040125AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068468.D

Aliphatic C9-C12	9682294.000	60.000	3.097	6.359	161371.567	162319.116	0.584
Aliphatic C12-C16	6682574.000	40.000	6.360	9.742	167064.350	168194.168	0.672
Aliphatic C16-C21	9959123.000	60.000	9.743	13.096	165985.383	171510.591	3.221
Aliphatic C21-C28	12456931.000	80.000	13.097	16.749	155711.638	168869.942	7.792
Aliphatic C28-C40	14862509.000	120.000	16.750	21.544	123854.242	136171.748	9.046
Aliphatic EPH	53643431.000	360.000	3.097	21.544	149009.531	157243.762	5.237