

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1675 SAS No.: Q1675 SDG No.: Q1675
Instrument ID: BNA_M Calibration Date(s): 03/13/2025 03/13/2025
Calibration Time(s): 09:02 13:37

LAB FILE ID:		RRF2.5 = BM049706.D		RRF005 = BM049707.D		RRF010 = BM049708.D		
		RRF020 = BM049709.D		RRF040 = BM049710.D		RRF050 = BM049711.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.027	1.092	1.201	1.176	1.239	1.185	8.1
Phenol-d6		1.325	1.383	1.498	1.470	1.553	1.488	6.9
Nitrobenzene-d5		0.343	0.361	0.392	0.385	0.403	0.390	7.6
Naphthalene		0.947	0.952	1.029	0.998	1.059	1.028	6.5
2-Methylnaphthalene		0.663	0.656	0.713	0.704	0.735	0.719	7.1
2-Fluorobiphenyl		1.388	1.419	1.563	1.587	1.696	1.597	9.6
2,4,6-Tribromophenol		0.173	0.193	0.226	0.250	0.270	0.234	19.2
Terphenyl-d14		0.978	1.021	1.204	1.207	1.352	1.187	12.6