

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01  
 Lab Code: CHEM Case No.: Q1675 SAS No.: Q1675 SDG No.: Q1675  
 Instrument ID: MSVOA\_Y Calibration Date(s): 03/27/2025 03/27/2025  
 Heated Purge: (Y/N) Y Calibration Time(s): 10:08 12:02  
 GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                   |  | RRF005 = VY021656.D |        | RRF010 = VY021657.D |        | RRF020 = VY021658.D |        | RRF050 = VY021659.D |       | RRF100 = VY021660.D |  | RRF150 = VY021661.D |  |
|--------------------------------|--|---------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|---------------------|--|---------------------|--|
| COMPOUND                       |  | RRF005              | RRF010 | RRF020              | RRF050 | RRF100              | RRF150 | RRF                 | % RSD |                     |  |                     |  |
| Dichlorodifluoromethane        |  | 0.602               | 0.577  | 0.493               | 0.506  | 0.485               | 0.447  | 0.518               | 11.4  |                     |  |                     |  |
| Chloromethane                  |  | 0.876               | 0.775  | 0.713               | 0.692  | 0.672               | 0.644  | 0.729               | 11.6  |                     |  |                     |  |
| Vinyl Chloride                 |  | 0.977               | 0.886  | 0.810               | 0.800  | 0.758               | 0.714  | 0.824               | 11.5  |                     |  |                     |  |
| Bromomethane                   |  | 0.709               | 0.613  | 0.537               | 0.513  | 0.494               | 0.506  | 0.562               | 14.9  |                     |  |                     |  |
| Chloroethane                   |  | 0.644               | 0.570  | 0.544               | 0.512  | 0.490               | 0.476  | 0.539               | 11.5  |                     |  |                     |  |
| Trichlorofluoromethane         |  | 1.212               | 1.124  | 1.104               | 1.013  | 0.967               | 0.914  | 1.056               | 10.5  |                     |  |                     |  |
| 1,1,2-Trichlorotrifluoroethane |  | 0.668               | 0.601  | 0.553               | 0.536  | 0.510               | 0.473  | 0.557               | 12.4  |                     |  |                     |  |
| 1,1-Dichloroethene             |  | 0.589               | 0.564  | 0.508               | 0.514  | 0.496               | 0.477  | 0.524               | 8.2   |                     |  |                     |  |
| Acetone                        |  | 0.145               | 0.123  | 0.102               | 0.107  | 0.087               | 0.102  | 0.111               | 18.3  |                     |  |                     |  |
| Carbon Disulfide               |  | 1.956               | 1.828  | 1.709               | 1.696  | 1.623               | 1.552  | 1.727               | 8.4   |                     |  |                     |  |
| Methyl tert-butyl Ether        |  | 1.380               | 1.335  | 1.247               | 1.325  | 1.256               | 1.292  | 1.306               | 3.9   |                     |  |                     |  |
| Methyl Acetate                 |  | 0.329               | 0.286  | 0.269               | 0.265  | 0.243               | 0.257  | 0.275               | 11    |                     |  |                     |  |
| Methylene Chloride             |  | 0.870               | 0.694  | 0.594               | 0.564  | 0.518               | 0.514  | 0.626               | 21.8  |                     |  |                     |  |
| trans-1,2-Dichloroethene       |  | 0.634               | 0.608  | 0.568               | 0.567  | 0.551               | 0.536  | 0.577               | 6.3   |                     |  |                     |  |
| 1,1-Dichloroethane             |  | 1.178               | 1.091  | 1.024               | 1.040  | 0.992               | 0.968  | 1.049               | 7.3   |                     |  |                     |  |
| Cyclohexane                    |  | 1.170               | 1.032  | 0.912               | 0.918  | 0.886               | 0.818  | 0.956               | 13.1  |                     |  |                     |  |
| 2-Butanone                     |  | 0.181               | 0.166  | 0.149               | 0.159  | 0.140               | 0.155  | 0.158               | 8.9   |                     |  |                     |  |
| Carbon Tetrachloride           |  | 0.637               | 0.594  | 0.546               | 0.564  | 0.547               | 0.507  | 0.566               | 7.9   |                     |  |                     |  |
| cis-1,2-Dichloroethene         |  | 0.700               | 0.658  | 0.626               | 0.652  | 0.631               | 0.629  | 0.649               | 4.3   |                     |  |                     |  |
| Bromochloromethane             |  | 0.513               | 0.470  | 0.424               | 0.426  | 0.420               | 0.402  | 0.442               | 9.3   |                     |  |                     |  |
| Chloroform                     |  | 1.235               | 1.141  | 1.062               | 1.081  | 1.026               | 1.000  | 1.091               | 7.9   |                     |  |                     |  |
| 1,1,1-Trichloroethane          |  | 1.155               | 1.013  | 0.959               | 0.967  | 0.929               | 0.893  | 0.986               | 9.3   |                     |  |                     |  |
| Methylcyclohexane              |  | 0.593               | 0.599  | 0.569               | 0.619  | 0.612               | 0.550  | 0.590               | 4.5   |                     |  |                     |  |
| Benzene                        |  | 1.606               | 1.527  | 1.429               | 1.499  | 1.445               | 1.371  | 1.479               | 5.6   |                     |  |                     |  |
| 1,2-Dichloroethane             |  | 0.467               | 0.445  | 0.399               | 0.420  | 0.392               | 0.379  | 0.417               | 8.1   |                     |  |                     |  |
| Trichloroethene                |  | 0.412               | 0.398  | 0.362               | 0.374  | 0.366               | 0.348  | 0.377               | 6.4   |                     |  |                     |  |
| 1,2-Dichloropropane            |  | 0.383               | 0.366  | 0.334               | 0.353  | 0.340               | 0.325  | 0.350               | 6.2   |                     |  |                     |  |
| Bromodichloromethane           |  | 0.571               | 0.546  | 0.504               | 0.525  | 0.508               | 0.484  | 0.523               | 6     |                     |  |                     |  |
| 4-Methyl-2-Pentanone           |  | 0.232               | 0.240  | 0.219               | 0.241  | 0.222               | 0.226  | 0.230               | 3.9   |                     |  |                     |  |
| Toluene                        |  | 0.940               | 0.943  | 0.893               | 0.951  | 0.928               | 0.885  | 0.923               | 3     |                     |  |                     |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                     |                  |                      |                   |
|---------------------|------------------|----------------------|-------------------|
| Lab Name:           | <u>CHEMTECH</u>  | Contract:            | <u>GENV01</u>     |
| Lab Code:           | <u>CHEM</u>      | Case No.:            | <u>Q1675</u>      |
| Instrument ID:      | <u>MSVOA_Y</u>   | SAS No.:             | <u>Q1675</u>      |
| Heated Purge: (Y/N) | <u>Y</u>         | SDG No.:             | <u>Q1675</u>      |
| GC Column:          | <u>RXI-624</u>   | Calibration Date(s): | <u>03/27/2025</u> |
| ID:                 | <u>0.25 (mm)</u> | Calibration Time(s): | <u>10:08</u>      |
|                     |                  |                      | <u>12:02</u>      |

|                             |        |                     |        |                     |        |                     |       |       |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                |        | RRF005 = VY021656.D |        | RRF010 = VY021657.D |        | RRF020 = VY021658.D |       |       |
|                             |        | RRF050 = VY021659.D |        | RRF100 = VY021660.D |        | RRF150 = VY021661.D |       |       |
| COMPOUND                    | RRF005 | RRF010              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |
| t-1,3-Dichloropropene       | 0.456  | 0.470               | 0.439  | 0.486               | 0.475  | 0.461               | 0.465 | 3.5   |
| cis-1,3-Dichloropropene     | 0.553  | 0.558               | 0.528  | 0.565               | 0.549  | 0.534               | 0.548 | 2.6   |
| 1,1,2-Trichloroethane       | 0.280  | 0.283               | 0.249  | 0.261               | 0.247  | 0.240               | 0.260 | 6.9   |
| 2-Hexanone                  | 0.154  | 0.154               | 0.144  | 0.164               | 0.148  | 0.154               | 0.153 | 4.2   |
| Dibromochloromethane        | 0.371  | 0.367               | 0.340  | 0.364               | 0.348  | 0.337               | 0.355 | 4.1   |
| 1,2-Dibromoethane           | 0.267  | 0.250               | 0.230  | 0.254               | 0.233  | 0.229               | 0.244 | 6.3   |
| Tetrachloroethene           | 0.534  | 0.503               | 0.453  | 0.449               | 0.412  | 0.382               | 0.455 | 12.3  |
| Chlorobenzene               | 1.244  | 1.178               | 1.110  | 1.155               | 1.116  | 1.073               | 1.146 | 5.3   |
| Ethyl Benzene               | 1.949  | 1.916               | 1.918  | 2.051               | 2.023  | 1.935               | 1.965 | 2.9   |
| m/p-Xylenes                 | 0.726  | 0.753               | 0.744  | 0.787               | 0.774  | 0.731               | 0.753 | 3.2   |
| o-Xylene                    | 0.665  | 0.670               | 0.676  | 0.738               | 0.725  | 0.694               | 0.695 | 4.4   |
| Styrene                     | 1.064  | 1.110               | 1.134  | 1.232               | 1.229  | 1.181               | 1.158 | 5.8   |
| Bromoform                   | 0.241  | 0.239               | 0.229  | 0.237               | 0.225  | 0.221               | 0.232 | 3.4   |
| Isopropylbenzene            | 3.628  | 3.701               | 3.582  | 3.826               | 3.843  | 3.702               | 3.714 | 2.8   |
| 1,1,2,2-Tetrachloroethane   | 0.764  | 0.720               | 0.616  | 0.657               | 0.619  | 0.631               | 0.668 | 9.1   |
| 1,3-Dichlorobenzene         | 1.895  | 1.774               | 1.707  | 1.748               | 1.728  | 1.689               | 1.757 | 4.2   |
| 1,4-Dichlorobenzene         | 1.876  | 1.794               | 1.704  | 1.741               | 1.673  | 1.626               | 1.736 | 5.2   |
| 1,2-Dichlorobenzene         | 1.598  | 1.607               | 1.496  | 1.524               | 1.486  | 1.465               | 1.529 | 3.9   |
| 1,2-Dibromo-3-Chloropropane | 0.112  | 0.111               | 0.097  | 0.102               | 0.096  | 0.102               | 0.103 | 6.6   |
| 1,2,4-Trichlorobenzene      | 0.771  | 0.781               | 0.787  | 0.878               | 0.870  | 0.889               | 0.829 | 6.6   |
| 1,2,3-Trichlorobenzene      | 0.653  | 0.681               | 0.673  | 0.746               | 0.737  | 0.753               | 0.707 | 6.1   |
| 1,2-Dichloroethane-d4       | 0.612  | 0.559               | 0.502  | 0.548               | 0.510  | 0.520               | 0.542 | 7.5   |
| Dibromofluoromethane        | 0.357  | 0.321               | 0.290  | 0.339               | 0.325  | 0.313               | 0.324 | 7     |
| Toluene-d8                  | 1.277  | 1.240               | 1.127  | 1.306               | 1.252  | 1.201               | 1.234 | 5.1   |
| 4-Bromofluorobenzene        | 0.447  | 0.410               | 0.378  | 0.439               | 0.424  | 0.407               | 0.417 | 6     |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.