

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM				
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM				
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m		
STD. NAME		STD REF.#					
Tune/Reschk		VP133499					
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508					
CCC		VP133510					
Internal Standard/PEM		VP131783					
ICV/I.BLK		VP133509					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021655.D	27 Mar 2025 09:23		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021656.D	27 Mar 2025 10:08		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY021657.D	27 Mar 2025 10:31		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021658.D	27 Mar 2025 10:54	Comp.#20 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021659.D	27 Mar 2025 11:16		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021660.D	27 Mar 2025 11:39		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021661.D	27 Mar 2025 12:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021662.D	27 Mar 2025 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY032725	VY021663.D	27 Mar 2025 13:15		SY/MD	Ok,M
10	VY0327SBL01	VY0327SBL01	VY021664.D	27 Mar 2025 13:59		SY/MD	Ok
11	VY0327SBS01	VY0327SBS01	VY021665.D	27 Mar 2025 14:38		SY/MD	Ok,M
12	VY0327SBSD01	VY0327SBSD01	VY021666.D	27 Mar 2025 15:00		SY/MD	Ok,M
13	Q1661-02	351	VY021667.D	27 Mar 2025 15:24	vial-B Not purge	SY/MD	Not Ok
14	Q1645-03RE	TP-4-VOCRE	VY021668.D	27 Mar 2025 15:48	vial-B Internal Standard Fail	SY/MD	Confirms
15	Q1662-03	TP-6-VOC	VY021669.D	27 Mar 2025 16:11	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1662-07	TP-7-VOC	VY021670.D	27 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q1661-02	351	VY021671.D	27 Mar 2025 16:58	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME	STD REF.#				
Tune/Reschk	VP133499				
Initial Calibration Stds	VP133500,VP133501,VP133502,VP133503,VP133505,VP133508				
CCC	VP133510				
Internal Standard/PEM	VP131783				
ICV/I.BLK	VP133509				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1650-02	STOCK-PILE-BIN2	VY021672.D	27 Mar 2025 17:21	vial-A NOT PURGE	SY/MD	Not Ok
19	Q1654-01	RT3407	VY021673.D	27 Mar 2025 17:45	vial-A Internal Standard Fail	SY/MD	Ok
20	Q1656-01	VNJ239	VY021674.D	27 Mar 2025 18:08	vial-A NOT PURGE	SY/MD	Not Ok
21	Q1648-09	TP-7	VY021675.D	27 Mar 2025 18:32	vial-A	SY/MD	Ok
22	Q1648-07	TP-6	VY021676.D	27 Mar 2025 18:55	vial-A	SY/MD	Ok
23	Q1648-05	TP-3	VY021677.D	27 Mar 2025 19:19	vial-A	SY/MD	Ok
24	Q1648-01	TP-4	VY021678.D	27 Mar 2025 19:42	vial-A	SY/MD	Ok
25	Q1648-03	TP-5	VY021679.D	27 Mar 2025 20:05	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021680.D	27 Mar 2025 20:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY033125

Review By	Amit Patel	Review On	4/1/2025 1:18:02 PM
Supervise By	Maresh Dadoda	Supervise On	4/1/2025 1:18:58 PM
SubDirectory	VY033125	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133540
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133538,VP133539 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021709.D	31 Mar 2025 09:37		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021710.D	31 Mar 2025 10:08		SY/MD	Ok,M
3	VY0331SBL01	VY0331SBL01	VY021711.D	31 Mar 2025 11:04		SY/MD	Ok
4	VY0331SBS01	VY0331SBS01	VY021712.D	31 Mar 2025 11:43		SY/MD	Ok,M
5	VY0331SBSD01	VY0331SBSD01	VY021713.D	31 Mar 2025 12:06		SY/MD	Ok,M
6	Q1657-01RE	72-11998RE	VY021714.D	31 Mar 2025 13:04	vial-B Internal Standard Fail	SY/MD	Confirms
7	Q1679-03	TP-10-VOC	VY021715.D	31 Mar 2025 13:27	vial-A	SY/MD	Ok
8	Q1679-07	TP-9-VOC	VY021716.D	31 Mar 2025 13:51	vial-A	SY/MD	Ok
9	Q1671-03	WC-1-VOC	VY021717.D	31 Mar 2025 14:14	vial-A	SY/MD	Ok
10	Q1675-01	P1	VY021718.D	31 Mar 2025 14:38	vial-A	SY/MD	Ok
11	Q1675-05	P5	VY021719.D	31 Mar 2025 15:01	vial-A	SY/MD	Ok
12	Q1675-06	P6	VY021720.D	31 Mar 2025 15:25	vial-A	SY/MD	Ok
13	Q1675-07	T1	VY021721.D	31 Mar 2025 15:48	vial-A	SY/MD	Ok
14	Q1675-08	T2	VY021722.D	31 Mar 2025 16:11	vial-A	SY/MD	Ok
15	Q1675-09	T3	VY021723.D	31 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
16	Q1675-10	T4	VY021724.D	31 Mar 2025 16:58	vial-A	SY/MD	Ok
17	Q1675-11	T5	VY021725.D	31 Mar 2025 17:22	vial-A	SY/MD	Ok
18	Q1672-03	TP-8-VOC	VY021726.D	31 Mar 2025 17:45		SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY033125

Review By	Amit Patel	Review On	4/1/2025 1:18:02 PM		
Supervise By	Mahesh Dadoda	Supervise On	4/1/2025 1:18:58 PM		
SubDirectory	VY033125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133540
CCC	VP133538,VP133539
Internal Standard/PEM	VP131783
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q1674-01	RT5358	VY021727.D	31 Mar 2025 18:08		SY/MD	Ok
20	Q1674-03	72-11991	VY021728.D	31 Mar 2025 18:32		SY/MD	Ok
21	VSTDCCC050	VSTDCCC050EC	VY021729.D	31 Mar 2025 18:55		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY040125

Review By	Amit Patel	Review On	4/2/2025 2:13:19 PM
Supervise By	Mahesh Dadoda	Supervise On	4/2/2025 2:15:14 PM
SubDirectory	VY040125	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133555
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133556,VP133557 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021730.D	01 Apr 2025 09:35		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021731.D	01 Apr 2025 10:13		SY/MD	Ok,M
3	VY0401SBL01	VY0401SBL01	VY021732.D	01 Apr 2025 10:52		SY/MD	Ok
4	VY0401SBS01	VY0401SBS01	VY021733.D	01 Apr 2025 11:30		SY/MD	Ok,M
5	VY0401SBSD01	VY0401SBSD01	VY021734.D	01 Apr 2025 11:52		SY/MD	Ok,M
6	Q1681-03	TP-12-VOC	VY021735.D	01 Apr 2025 12:36	vial-A	SY/MD	Ok
7	Q1680-03	TP-4-VOC	VY021736.D	01 Apr 2025 12:59	vial-A	SY/MD	Ok
8	Q1675-09	T3	VY021737.D	01 Apr 2025 13:22	vial-B	SY/MD	Ok
9	Q1675-12	T6	VY021738.D	01 Apr 2025 13:46	vial-A	SY/MD	Ok
10	IBLK	IBLK	VY021739.D	01 Apr 2025 14:19		SY/MD	Ok
11	Q1664-02MS	P001-BBDGA-001-01M	VY021740.D	01 Apr 2025 14:42	vial-B Internal Standard Fail	SY/MD	Ok,M
12	Q1664-03MSD	P001-BBDGA-001-01M	VY021741.D	01 Apr 2025 15:05	vial-B	SY/MD	Ok,M
13	IBLK	IBLK	VY021742.D	01 Apr 2025 15:28		SY/MD	Ok
14	Q1687-01	72-12016	VY021743.D	01 Apr 2025 15:52	Not purged	SY/MD	Not Ok
15	Q1688-01	OR-02-040125	VY021744.D	01 Apr 2025 16:15	vial-A Not purge	SY/MD	Not Ok
16	Q1692-01	NB-08-04012025	VY021745.D	01 Apr 2025 16:38	vial-A Internal Standard Fail	SY/MD	ReRun
17	VSTDCCC050	VSTDCCC050EC	VY021746.D	01 Apr 2025 17:01		SY/MD	Ok,M

M : Manual Integration