

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1675 SAS No.: Q1675 SDG NO.: Q1675
 Lab File ID: VY021710.D Date Analyzed: 03/31/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	152394	13.347				
UPPER LIMIT	304788	13.847				
LOWER LIMIT	76197	12.847				
EPA SAMPLE NO.						
P1	151796	13.35				
P5	108429	13.35				
P6	189318	13.35				
T1	185708	13.35				
T2	256507	13.35				
T4	250658	13.35				
T5	244186	13.35				
VY0331SBL01	140307	13.35				
VY0331SBS01	145397	13.35				
VY0331SBSD01	150013	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.