

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070118.D Time Analyzed: 11:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.494	4.386	4.586	230024400.000	220389000.000	-4.2
Aroclor-1016-2	4.513	4.404	4.604	387085800.000	310701000.000	-19.7
Aroclor-1016-3	4.571	4.463	4.663	252199800.000	227746000.000	-9.7
Aroclor-1016-4	4.661	4.551	4.751	199753500.000	175670000.000	-12.1
Aroclor-1016-5	4.945	4.836	5.036	196065600.000	176518000.000	-10.0
Aroclor-1260-1	6.035	5.926	6.126	342744100.000	308864000.000	-9.9
Aroclor-1260-2	6.290	6.181	6.381	375376700.000	311223000.000	-17.1
Aroclor-1260-3	6.641	6.533	6.733	301270200.000	267554000.000	-11.2
Aroclor-1260-4	6.855	6.746	6.946	326357800.000	284736000.000	-12.8
Aroclor-1260-5	7.161	7.052	7.252	583624600.000	478047000.000	-18.1
DCB	8.512	8.403	8.603	3797495000.000	3543170000.000	-6.7
TCX	3.414	3.305	3.505	8939766000.000	8179750000.000	-8.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070118.D Time Analyzed: 11:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.785	3.687	3.887	223585300.000	188460000.000	-15.7
Aroclor-1016-2	3.801	3.702	3.902	365003900.000	373786000.000	2.4
Aroclor-1016-3	3.962	3.863	4.063	206649600.000	188666000.000	-8.7
Aroclor-1016-4	4.010	3.911	4.111	189703100.000	180586000.000	-4.8
Aroclor-1016-5	4.205	4.106	4.306	222979400.000	206455000.000	-7.4
Aroclor-1260-1	5.200	5.100	5.300	400599300.000	368329000.000	-8.1
Aroclor-1260-2	5.389	5.289	5.489	466852700.000	412992000.000	-11.5
Aroclor-1260-3	5.530	5.431	5.631	444827800.000	392023000.000	-11.9
Aroclor-1260-4	5.991	5.892	6.092	392153300.000	347573000.000	-11.4
Aroclor-1260-5	6.236	6.136	6.336	828820000.000	712917000.000	-14.0
DCB	7.548	7.449	7.649	5861421000.000	5442260000.000	-7.2
TCX	2.783	2.683	2.883	8065420000.000	7556320000.000	-6.3