

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070122.D Time Analyzed: 12:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.486	4.386	4.586	230024400.000	205406000.000	-10.7
Aroclor-1016-2	4.505	4.404	4.604	387085800.000	352002000.000	-9.1
Aroclor-1016-3	4.562	4.463	4.663	252199800.000	229717000.000	-8.9
Aroclor-1016-4	4.652	4.551	4.751	199753500.000	180634000.000	-9.6
Aroclor-1016-5	4.936	4.836	5.036	196065600.000	178730000.000	-8.8
Aroclor-1260-1	6.026	5.926	6.126	342744100.000	310535000.000	-9.4
Aroclor-1260-2	6.282	6.181	6.381	375376700.000	337323000.000	-10.1
Aroclor-1260-3	6.633	6.533	6.733	301270200.000	276646000.000	-8.2
Aroclor-1260-4	6.846	6.746	6.946	326357800.000	295885000.000	-9.3
Aroclor-1260-5	7.151	7.052	7.252	583624600.000	520361000.000	-10.8
DCB	8.502	8.403	8.603	3797495000.000	3570370000.000	-6.0
TCX	3.405	3.305	3.505	8939766000.000	8214370000.000	-8.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070122.D Time Analyzed: 12:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.786	3.687	3.887	223585300.000	211055000.000	-5.6
Aroclor-1016-2	3.801	3.702	3.902	365003900.000	347057000.000	-4.9
Aroclor-1016-3	3.962	3.863	4.063	206649600.000	193870000.000	-6.2
Aroclor-1016-4	4.010	3.911	4.111	189703100.000	180158000.000	-5.0
Aroclor-1016-5	4.205	4.106	4.306	222979400.000	210719000.000	-5.5
Aroclor-1260-1	5.199	5.100	5.300	400599300.000	380267000.000	-5.1
Aroclor-1260-2	5.388	5.289	5.489	466852700.000	442810000.000	-5.2
Aroclor-1260-3	5.529	5.431	5.631	444827800.000	418369000.000	-5.9
Aroclor-1260-4	5.990	5.892	6.092	392153300.000	371003000.000	-5.4
Aroclor-1260-5	6.235	6.136	6.336	828820000.000	779445000.000	-6.0
DCB	7.547	7.449	7.649	5861421000.000	5681090000.000	-3.1
TCX	2.784	2.683	2.883	8065420000.000	7706770000.000	-4.5