

Surrogate Summary

SDG No.: Q1676

Client: Chemtech Consulting Group

Analytical Method: SFAM_PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PQ070047.D	PIBLK-PQ070047.D	TCX	1	20	0.18	91		30	150
		DCB	1	40	0.38	94		30	150
		TCX	2	20	0.19	97		30	150
		DCB	2	40	0.39	98		30	150
I.BLK-PQ070117.D	PIBLK-PQ070117.D	TCX	1	20	0.18	92		30	150
		DCB	1	40	0.39	97		30	150
		TCX	2	20	0.20	101		30	150
		DCB	2	40	0.40	100		30	150
Q1676-13	PCB-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
Q1676-14	PCB-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
I.BLK-PQ070121.D	PIBLK-PQ070121.D	TCX	1	20	0.19	94		30	150
		DCB	1	40	0.40	101		30	150
		TCX	2	20	0.20	102		30	150
		DCB	2	40	0.41	103		30	150