

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/21/2025 03/21/2025

Client Sample No.: CCAL01 Date Analyzed: 04/01/2025

Lab Sample No.: INDA324 Data File : PD088052.D Time Analyzed: 13:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.709	6.639	6.779	2577180.000	2493100.000	-3.3
4,4'-DDT	7.025	6.955	7.095	2701560.000	2639900.000	-2.3
alpha-BHC	4.002	3.952	4.052	4110160.000	4037420.000	-1.8
DCB	9.078	8.978	9.178	3351440.000	3258370.000	-2.8
Dieldrin	6.351	6.281	6.421	3672580.000	3619950.000	-1.4
Endosulfan I	6.077	6.007	6.147	3344420.000	3306270.000	-1.1
Endrin	6.578	6.508	6.648	3065760.000	2992750.000	-2.4
gamma-BHC (Lindane)	4.333	4.283	4.383	3987070.000	3917630.000	-1.7
Heptachlor	4.933	4.883	4.983	4090350.000	4025840.000	-1.6
Methoxychlor	7.496	7.426	7.566	1489960.000	1460170.000	-2.0
TCX	3.552	3.502	3.602	1972170.000	1985160.000	0.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/21/2025 03/21/2025

Client Sample No.: CCAL01 Date Analyzed: 04/01/2025

Lab Sample No.: INDA324 Data File : PD088052.D Time Analyzed: 13:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	5.934	5.864	6.004	16337800.000	15487900.000	-5.2
4,4'-DDT	6.189	6.119	6.259	16751800.000	16241000.000	-3.0
alpha-BHC	3.397	3.347	3.447	24236900.000	23262900.000	-4.0
DCB	8.079	7.979	8.179	17755800.000	16522700.000	-6.9
Dieldrin	5.517	5.447	5.587	20235100.000	19352100.000	-4.4
Endosulfan I	5.251	5.181	5.321	17350700.000	18304100.000	5.5
Endrin	5.794	5.724	5.864	18150200.000	16978400.000	-6.5
gamma-BHC (Lindane)	3.733	3.683	3.783	22254300.000	21356900.000	-4.0
Heptachlor	4.087	4.037	4.137	22826800.000	21833900.000	-4.3
Methoxychlor	6.760	6.690	6.830	7450680.000	6923760.000	-7.1
TCX	2.883	2.833	2.933	15371300.000	14763700.000	-4.0