

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676
 Lab File ID: VX045559.D Date Analyzed: 04/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	95221	5.54	162684	6.76	143783	10.05
UPPER LIMIT	190442	6.044	325368	7.257	287566	10.549
LOWER LIMIT	47610.5	5.044	81342	6.257	71891.5	9.549
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	65833	5.55	129421	6.76	121603	10.05
Storage-Blank-WATER-REF-5	65062	5.54	128448	6.76	120205	10.05
Storage-Blank-WATER-REF-4	66549	5.54	128935	6.76	120457	10.05
Storage-Blank-SAMPLE-MANAGEMENT	65504	5.55	130677	6.76	123569	10.06
VX0403WBL01	66309	5.55	126206	6.76	118586	10.05
VX0403WBS01	95034	5.54	168178	6.76	149052	10.05
VX0403WBSD01	95765	5.55	168463	6.76	148868	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.