

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1676 SAS No.: Q1676 SDG NO.: Q1676
 Lab File ID: VX045559.D Date Analyzed: 04/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:30
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66270	12.018				
UPPER LIMIT	132540	12.518				
LOWER LIMIT	33135	11.518				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	51026	12.02				
Storage-Blank-WATER-REF-5	48761	12.02				
Storage-Blank-WATER-REF-4	48469	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	52066	12.02				
VX0403WBL01	50130	12.02				
VX0403WBS01	68919	12.02				
VX0403WBSD01	66417	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.