

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1689

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB135309

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.094	0.099	95	85-115	04/04/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	04/04/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	04/04/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1689

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB135317

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.0	7	100	90-110	04/01/2025
Sample ID:	CCV1						
Field pH		pH	7.0	7.0	100	90-110	04/01/2025
Sample ID:	CCV2						
Field pH		pH	7.0	7.0	100	90-110	04/01/2025

Initial and Continuing Calibration Verification

Client: METEM A GE POWER Business

SDG No.: Q1689

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB135317

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------