

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q1689

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB135309

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	04/04/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	04/04/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	04/04/2025

Initial and Continuing Calibration Blank Summary

Client: METEM A GE POWER Business

SDG No.: Q1689

Project: Parsippany Wastewater Quarterly 2025

RunNo.: LB135309

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------