

Duplicate Sample Summary

Client: METEM A GE POWER Business	SDG No.: Q1689
Project: Parsippany Wastewater Quarterly 2025	Sample ID: Q1689-01
Client ID: Q2 WASTEWATERDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0012	U	0.0012	U	1	0		04/04/2025
Field pH	pH	+/-20	7.69		7.70		1	0.13		04/01/2025

Duplicate Sample Summary

Client: METEM A GE POWER Business Project: Parsippany Wastewater Quarterly 2025 Client ID: Q2 WASTEWATERMSD	SDG No.: Q1689 Sample ID: Q1689-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.039		0.039		1	0		04/04/2025