

### Laboratory Control Sample Summary

**Client:** METEM A GE POWER Business

**SDG No.:** Q1689

**Project:** Parsippany Wastewater Quarterly 2025

**Run No.:** LB135309

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB167464BS |            |        |                 |            |                 |                     |               |
| Cyanide   | mg/L       | 0.1        | 0.094  |                 | 94         | 1               | 85-115              | 04/04/2025    |