



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1689

Contract: METE01

Lab Code: CHEM

Case No.: Q1689

SAS No.: Q1689

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:15	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	12:15	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:15	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	12:15	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	12:15	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1689

Contract: METE01

Lab Code: CHEM

Case No.: Q1689

SAS No.: Q1689

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:47	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	12:47	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:47	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	12:47	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	12:47	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	13:57	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	13:57	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	13:57	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	13:57	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	13:57	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
CCB03	Cadmium	0.26	+/-6.00	J	6.00	P	04/03/2025	14:43	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	14:43	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	14:43	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	14:43	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	14:43	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	15:36	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	15:36	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	15:36	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	15:36	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	15:36	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
CCB05	Cadmium	0.38	+/-6.00	J	6.00	P	04/03/2025	16:40	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	16:40	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	16:40	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	16:40	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	16:40	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	18:00	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	18:00	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:00	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	18:00	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business		SDG No.: Q1689							
Contract: METE01	Lab Code: CHEM	Case No.: Q1689	SAS No.: Q1689						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	18:00	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
CCB07	Cadmium	0.23	+/-6.00	J	6.00	P	04/03/2025	18:33	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	18:33	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:33	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	18:33	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	18:33	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>METEM A GE POWER Business</u>		SDG No.: <u>Q1689</u>							
Contract: <u>METE01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1689</u>	SAS No.: <u>Q1689</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/09/2025	14:30	LB135361
	Chromium	10.0	+/-10.0	U	10.0	P	04/09/2025	14:30	LB135361
	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	14:30	LB135361
	Lead	12.0	+/-12.0	U	12.0	P	04/09/2025	14:30	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	14:30	LB135361
	Silver	10.0	+/-10.0	U	10.0	P	04/09/2025	14:30	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	14:30	LB135361

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1689

Contract: METE01

Lab Code: CHEM

Case No.: Q1689

SAS No.: Q1689

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.29	+/-6.00	J	6.00	P	04/09/2025	15:00	LB135361
	Chromium	10.0	+/-10.0	U	10.0	P	04/09/2025	15:00	LB135361
	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	15:00	LB135361
	Lead	12.0	+/-12.0	U	12.0	P	04/09/2025	15:00	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	15:00	LB135361
	Silver	10.0	+/-10.0	U	10.0	P	04/09/2025	15:00	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	15:00	LB135361
CCB02	Cadmium	0.79	+/-6.00	J	6.00	P	04/09/2025	15:27	LB135361
	Chromium	10.0	+/-10.0	U	10.0	P	04/09/2025	15:27	LB135361
	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	15:27	LB135361
	Lead	12.0	+/-12.0	U	12.0	P	04/09/2025	15:27	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	15:27	LB135361
	Silver	10.0	+/-10.0	U	10.0	P	04/09/2025	15:27	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	15:27	LB135361
CCB03	Cadmium	0.59	+/-6.00	J	6.00	P	04/09/2025	16:30	LB135361
	Chromium	10.0	+/-10.0	U	10.0	P	04/09/2025	16:30	LB135361
	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	16:30	LB135361
	Lead	12.0	+/-12.0	U	12.0	P	04/09/2025	16:30	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	16:30	LB135361
	Silver	10.0	+/-10.0	U	10.0	P	04/09/2025	16:30	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	16:30	LB135361
CCB04	Cadmium	0.58	+/-6.00	J	6.00	P	04/09/2025	17:27	LB135361
	Chromium	10.0	+/-10.0	U	10.0	P	04/09/2025	17:27	LB135361
	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	17:27	LB135361
	Lead	12.0	+/-12.0	U	12.0	P	04/09/2025	17:27	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	17:27	LB135361
	Silver	10.0	+/-10.0	U	10.0	P	04/09/2025	17:27	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	17:27	LB135361
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	04/09/2025	18:13	LB135361
	Chromium	10.0	+/-10.0	U	10.0	P	04/09/2025	18:13	LB135361
	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	18:13	LB135361
	Lead	12.0	+/-12.0	U	12.0	P	04/09/2025	18:13	LB135361
	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	18:13	LB135361
	Silver	10.0	+/-10.0	U	10.0	P	04/09/2025	18:13	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	18:13	LB135361
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	04/09/2025	18:40	LB135361
	Chromium	10.0	+/-10.0	U	10.0	P	04/09/2025	18:40	LB135361
	Copper	20.0	+/-20.0	U	20.0	P	04/09/2025	18:40	LB135361
	Lead	12.0	+/-12.0	U	12.0	P	04/09/2025	18:40	LB135361

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1689

Contract: METE01

Lab Code: CHEM

Case No.: Q1689

SAS No.: Q1689

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Nickel	40.0	+/-40.0	U	40.0	P	04/09/2025	18:40	LB135361
	Silver	10.0	+/-10.0	U	10.0	P	04/09/2025	18:40	LB135361
	Zinc	40.0	+/-40.0	U	40.0	P	04/09/2025	18:40	LB135361