

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: METEM A GE POWER Business **SDG No.:** Q1689
Contract: METE01 **Lab Code:** CHEM **Case No.:** Q1689 **SAS No.:** Q1689
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-1.36	1.0	136	-5	7	04/03/2025	12:24	LB135289
	Chromium	56.8	52.0	109	42	62	04/03/2025	12:24	LB135289
	Copper	3.95	2.0	198	-18	22	04/03/2025	12:24	LB135289
	Lead	4.36			-12	12	04/03/2025	12:24	LB135289
	Nickel	1.10	2.0	55	-38	42	04/03/2025	12:24	LB135289
	Silver	0.24			-10	10	04/03/2025	12:24	LB135289
	Zinc	5.44			-40	40	04/03/2025	12:24	LB135289
ICSAB01	Cadmium	1000	972	103	826	1120	04/03/2025	12:28	LB135289
	Chromium	573	542	106	460	624	04/03/2025	12:28	LB135289
	Copper	512	511	100	434	588	04/03/2025	12:28	LB135289
	Lead	55.2	49.0	113	37	61	04/03/2025	12:28	LB135289
	Nickel	1010	954	106	810	1100	04/03/2025	12:28	LB135289
	Silver	191	201	95	170	232	04/03/2025	12:28	LB135289
	Zinc	1070	952	112	809	1095	04/03/2025	12:28	LB135289
ICSA	Cadmium	-4.04	1.0	404	-5	7	04/03/2025	12:33	LB135289
	Chromium	52.2	52.0	100	42	62	04/03/2025	12:33	LB135289
	Copper	10.4	2.0	522	-18	22	04/03/2025	12:33	LB135289
	Lead	8.69			-12	12	04/03/2025	12:33	LB135289
	Nickel	-1.42	2.0	71	-38	42	04/03/2025	12:33	LB135289
	Silver	4.98			-10	10	04/03/2025	12:33	LB135289
	Zinc	-5.86			-40	40	04/03/2025	12:33	LB135289
ICSAB	Cadmium	997	972	103	826	1120	04/03/2025	12:37	LB135289
	Chromium	596	542	110	460	624	04/03/2025	12:37	LB135289
	Copper	538	511	105	434	588	04/03/2025	12:37	LB135289
	Lead	56.0	49.0	114	37	61	04/03/2025	12:37	LB135289
	Nickel	1010	954	106	810	1100	04/03/2025	12:37	LB135289
	Silver	191	201	95	170	232	04/03/2025	12:37	LB135289
	Zinc	1070	952	112	809	1095	04/03/2025	12:37	LB135289
ICSA01	Cadmium	-1.80	1.0	180	-5	7	04/09/2025	14:39	LB135361
	Chromium	54.6	52.0	105	42	62	04/09/2025	14:39	LB135361
	Copper	0.93	2.0	46	-18	22	04/09/2025	14:39	LB135361
	Lead	6.82			-12	12	04/09/2025	14:39	LB135361
	Nickel	2.38	2.0	119	-38	42	04/09/2025	14:39	LB135361
	Silver	0.79			-10	10	04/09/2025	14:39	LB135361
	Zinc	2.22			-40	40	04/09/2025	14:39	LB135361
ICSAB01	Cadmium	975	972	100	826	1120	04/09/2025	14:43	LB135361
	Chromium	547	542	101	460	624	04/09/2025	14:43	LB135361
	Copper	483	511	94	434	588	04/09/2025	14:43	LB135361
	Lead	51.3	49.0	105	37	61	04/09/2025	14:43	LB135361
	Nickel	982	954	103	810	1100	04/09/2025	14:43	LB135361
	Silver	186	201	92	170	232	04/09/2025	14:43	LB135361

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>METEM A GE POWER Business</u>	SDG No.: <u>Q1689</u>
Contract: <u>METE01</u> Lab Code: <u>CHEM</u>	Case No.: <u>Q1689</u> SAS No.: <u>Q1689</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Zinc	1030	952	108	809	1095	04/09/2025	14:43	LB135361
ICSA	Cadmium	-2.52	1.0	252	-5	7	04/09/2025	14:48	LB135361
	Chromium	53.0	52.0	102	42	62	04/09/2025	14:48	LB135361
	Copper	-2.16	2.0	108	-18	22	04/09/2025	14:48	LB135361
	Lead	16.4			-12	12	04/09/2025	14:48	LB135361
	Nickel	8.09	2.0	404	-38	42	04/09/2025	14:48	LB135361
	Silver	-2.28			-10	10	04/09/2025	14:48	LB135361
	Zinc	9.91			-40	40	04/09/2025	14:48	LB135361
ICSAB	Cadmium	1000	972	103	826	1120	04/09/2025	14:52	LB135361
	Chromium	574	542	106	460	624	04/09/2025	14:52	LB135361
	Copper	527	511	103	434	588	04/09/2025	14:52	LB135361
	Lead	52.0	49.0	106	37	61	04/09/2025	14:52	LB135361
	Nickel	1010	954	106	810	1100	04/09/2025	14:52	LB135361
	Silver	180	201	90	170	232	04/09/2025	14:52	LB135361
	Zinc	1030	952	108	809	1095	04/09/2025	14:52	LB135361