

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: METEM A GE POWER Business **Level:** LOW **SDG No.:** Q1689
Contract: METE01 **Lab Code:** CHEM **Case No.:** Q1689 **SAS No.:** Q1689
Matrix: Water **Sample ID:** Q1689-01 **Client ID:** Q2 WASTEWATERDUP
Percent Solids for Sample: NA **Duplicate ID** Q1689-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	30.0	UD	30.0	UD	200		P
Chromium	ug/L	20	126	D	124	D	2		P
Copper	ug/L	20	100	UD	100	UD	200		P
Lead	ug/L	20	60.0	UD	60.0	UD	200		P
Nickel	ug/L	20	1450	D	1440	D	1		P
Silver	ug/L	20	50.0	UD	50.0	UD	200		P
Zinc	ug/L	20	86.2	JD	84.4	JD	2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: METEM A GE POWER Business **Level:** LOW **SDG No.:** Q1689
Contract: METE01 **Lab Code:** CHEM **Case No.:** Q1689 **SAS No.:** Q1689
Matrix: Water **Sample ID:** Q1689-01MS **Client ID:** Q2 WASTEWATERMSD
Percent Solids for Sample: NA **Duplicate ID** Q1689-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	93.9	D	95.3	D	1		P
Chromium	ug/L	20	296	D	298	D	1		P
Copper	ug/L	20	129	D	131	D	2		P
Lead	ug/L	20	423	D	426	D	1		P
Nickel	ug/L	20	1630	D	1640	D	1		P
Silver	ug/L	20	49.4	JD	50.2	D	2		P
Zinc	ug/L	20	204	D	201	D	1		P