

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: METEM A GE POWER Business

SDG No.: Q1689

Contract: METE01

Lab Code: CHEM

Case No.: Q1689

SAS No.: Q1689

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167442BS							
Cadmium	ug/L	100	97.2		97	85 - 115	P
Chromium	ug/L	200	206		103	85 - 115	P
Copper	ug/L	150	158		105	85 - 115	P
Lead	ug/L	500	478		96	85 - 115	P
Nickel	ug/L	250	249		100	85 - 115	P
Silver	ug/L	37.5	37.9		101	85 - 115	P
Zinc	ug/L	100	103		103	85 - 115	P