

LB135269


7470A

INSTRUMENT ID: CV1

Linear

$\mu$  Abs.:

39587



A= 0.0000e+000

B= 2.5692e-004

C= 2.8351e-002

Rho= 0.9989419

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.046  | 0.046  | 68    | 0.000      | 68    | 0     |       |       |       | 17 |
| 0.2    | 0.200  | 0.233  | 0.033  | 798   | 0.0 %      | 798   |       |       |       |       | -1 |
| 2.5    | 2.500  | 2.466  | -0.034 | 9490  | 0.0 %      | 9490  |       |       |       |       | 2  |
| 5.0    | 5.000  | 5.100  | 0.100  | 19741 | 0.0 %      | 19741 |       |       |       |       | -5 |
| 7.5    | 7.500  | 7.155  | -0.345 | 27740 | 0.0 %      | 27740 |       |       |       |       | 2  |
| 10.0   | 10.000 | 10.199 | 0.199  | 39587 | 0.0 %      | 39587 |       |       |       |       |    |

# LB135269 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Apr 2025 13:29:47

| Sample ID       | Extended ID | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50          | 68     | -       | 0.0000   | 7470A  | PPB   | 02 Apr 2025 13:40:28 | S    | Std  |
| 0.2 - 1         | 50.2        | 798    | -       | 0.2000   | 7470A  | PPB   | 02 Apr 2025 13:42:45 | S    | Std  |
| 2.5 - 1         | 52.5        | 9490   | -       | 2.5000   | 7470A  | PPB   | 02 Apr 2025 13:45:02 | S    | Std  |
| 5.0 - 1         | 55          | 19741  | -       | 5.0000   | 7470A  | PPB   | 02 Apr 2025 13:47:19 | S    | Std  |
| 7.5 - 1         | 57.5        | 27740  | -       | 7.5000   | 7470A  | PPB   | 02 Apr 2025 13:55:02 | S    | Std  |
| 10.0 - 1        | 510         | 39587  | -       | 10.0000  | 7470A  | PPB   | 02 Apr 2025 14:00:04 | S    | Std  |
| ICV73 - 1       | ICV73       | 15602  | 4.0368  | -        | 7470A  | PPB   | 02 Apr 2025 14:10:13 | U    | SMPL |
| ICB73 - 1       | ICB73       | -215   | -0.0269 | -        | 7470A  | PPB   | 02 Apr 2025 14:12:29 | U    | SMPL |
| CCV23 - 1       | CCV23       | 18496  | 4.7803  | -        | 7470A  | PPB   | 02 Apr 2025 14:14:46 | U    | SMPL |
| CCB23 - 1       | CCB23       | -229   | -0.0305 | -        | 7470A  | PPB   | 02 Apr 2025 14:17:01 | U    | SMPL |
| CRA - 1         | CRA         | 676    | 0.2020  | -        | 7470A  | PPB   | 02 Apr 2025 14:19:18 | U    | SMPL |
| HighStd - 1     | HighStd     | 37584  | 9.6843  | -        | 7470A  | PPB   | 02 Apr 2025 14:21:33 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 23255  | 6.0030  | -        | 7470A  | PPB   | 02 Apr 2025 14:23:49 | U    | SMPL |
| PB167425BL - 1  | PBW         | -362   | -0.0647 | -        | 7470A  | PPB   | 02 Apr 2025 14:26:07 | U    | SMPL |
| Q1666-01 - 1    | TW-WTS-05   | -351   | -0.0618 | -        | 7470A  | PPB   | 02 Apr 2025 14:30:42 | U    | SMPL |
| Q1690-01 - 1    | MW112010    | 94     | 0.0525  | -        | 7470A  | PPB   | 02 Apr 2025 14:33:00 | U    | SMPL |
| Q1690-01DUP - 1 | MW112010DUP | -10    | 0.0258  | -        | 7470A  | PPB   | 02 Apr 2025 14:35:17 | U    | SMPL |
| Q1690-01MS - 1  | MW112010MS  | 16137  | 4.1742  | -        | 7470A  | PPB   | 02 Apr 2025 14:40:05 | U    | SMPL |
| Q1690-01MSD - 1 | MW112010MSD | 15761  | 4.0776  | -        | 7470A  | PPB   | 02 Apr 2025 14:42:19 | U    | SMPL |
| CCV24 - 1       | CCV24       | 18250  | 4.7171  | -        | 7470A  | PPB   | 02 Apr 2025 14:44:37 | U    | SMPL |
| CCB24 - 1       | CCB24       | -258   | -0.0379 | -        | 7470A  | PPB   | 02 Apr 2025 14:46:55 | U    | SMPL |
| Q1690-01LX5 - 1 |             | -5     | 0.0271  | -        | 7470A  | PPB   | 02 Apr 2025 14:54:11 | U    | SMPL |
| Q1690-01A - 1   |             | 14334  | 3.7110  | -        | 7470A  | PPB   | 02 Apr 2025 14:56:26 | U    | SMPL |
| PB167425BS - 1  | LCSW        | 13207  | 3.4215  | -        | 7470A  | PPB   | 02 Apr 2025 14:58:41 | U    | SMPL |
| CCV25 - 1       | CCV25       | 17494  | 4.5229  | -        | 7470A  | PPB   | 02 Apr 2025 15:03:33 | U    | SMPL |
| CCB25 - 1       | CCB25       | -281   | -0.0438 | -        | 7470A  | PPB   | 02 Apr 2025 15:05:50 | U    | SMPL |