



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental SDG No.: Q1690
Contract: GENV01 Lab Code: CHEM Case No.: Q1690 SAS No.: Q1690
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	101	40 - 160	CV	04/02/2025	14:19	LB135269
CRI01	Aluminum	108	100	108	40 - 160	P	04/03/2025	12:20	LB135289
	Antimony	52.6	50.0	105	40 - 160	P	04/03/2025	12:20	LB135289
	Arsenic	21.0	20.0	105	40 - 160	P	04/03/2025	12:20	LB135289
	Barium	82.8	100	83	40 - 160	P	04/03/2025	12:20	LB135289
	Beryllium	5.91	6.0	98	40 - 160	P	04/03/2025	12:20	LB135289
	Cadmium	5.76	6.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Calcium	1940	2000	97	40 - 160	P	04/03/2025	12:20	LB135289
	Chromium	10.00	10.0	100	40 - 160	P	04/03/2025	12:20	LB135289
	Cobalt	28.9	30.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Copper	22.8	20.0	114	40 - 160	P	04/03/2025	12:20	LB135289
	Iron	110	100	110	40 - 160	P	04/03/2025	12:20	LB135289
	Lead	11.6	12.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Magnesium	2020	2000	101	40 - 160	P	04/03/2025	12:20	LB135289
	Manganese	16.4	20.0	82	40 - 160	P	04/03/2025	12:20	LB135289
	Nickel	38.7	40.0	97	40 - 160	P	04/03/2025	12:20	LB135289
	Potassium	1850	2000	92	40 - 160	P	04/03/2025	12:20	LB135289
	Selenium	20.4	20.0	102	40 - 160	P	04/03/2025	12:20	LB135289
	Silver	10.7	10.0	107	40 - 160	P	04/03/2025	12:20	LB135289
	Sodium	1950	2000	97	40 - 160	P	04/03/2025	12:20	LB135289
	Thallium	41.1	40.0	103	40 - 160	P	04/03/2025	12:20	LB135289
	Vanadium	38.5	40.0	96	40 - 160	P	04/03/2025	12:20	LB135289
	Zinc	42.8	40.0	107	40 - 160	P	04/03/2025	12:20	LB135289