



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental				SDG No.:	Q1690				
Contract:	GENV01	Lab Code:	CHEM		Case No.:	Q1690	SAS No.:	Q1690		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB73	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	14:12	LB135269	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1690

Contract: GENV01

Lab Code: CHEM

Case No.: Q1690

SAS No.: Q1690

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB23	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	14:17	LB135269
CCB24	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	14:46	LB135269
CCB25	Mercury	0.20	+/-0.20	U	0.20	CV	04/02/2025	15:05	LB135269

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1690							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1690	SAS No.: Q1690						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	04/03/2025	12:15	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	12:15	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	12:15	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	12:15	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:15	LB135289
	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:15	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	12:15	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	12:15	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	12:15	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:15	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	12:15	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	12:15	LB135289
	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	12:15	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	12:15	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	12:15	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	12:15	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	12:15	LB135289
	Sodium	2000	+/-2000	U	2000	P	04/03/2025	12:15	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:15	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1690</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1690</u>	SAS No.: <u>Q1690</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	04/03/2025	12:47	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	12:47	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	12:47	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	12:47	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:47	LB135289
	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	12:47	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	12:47	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	12:47	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	12:47	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	12:47	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	12:47	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	12:47	LB135289
	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	12:47	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	12:47	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	12:47	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	12:47	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	12:47	LB135289
	Sodium	2000	+/-2000	U	2000	P	04/03/2025	12:47	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	12:47	LB135289
CCB02	Aluminum	100	+/-100	U	100	P	04/03/2025	13:57	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	13:57	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	13:57	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	13:57	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	13:57	LB135289
	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	13:57	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	13:57	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	13:57	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	13:57	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	13:57	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	13:57	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	13:57	LB135289
	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	13:57	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	13:57	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	13:57	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	13:57	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1690</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1690</u>	SAS No.: <u>Q1690</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	13:57	LB135289
	Sodium	2000	+/-2000	U	2000	P	04/03/2025	13:57	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	13:57	LB135289
CCB03	Aluminum	100	+/-100	U	100	P	04/03/2025	14:43	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	14:43	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	14:43	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	14:43	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	14:43	LB135289
	Cadmium	0.26	+/-6.00	J	6.00	P	04/03/2025	14:43	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	14:43	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	14:43	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	14:43	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	14:43	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	14:43	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	14:43	LB135289
	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	14:43	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	14:43	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	14:43	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	14:43	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	14:43	LB135289
	Sodium	2000	+/-2000	U	2000	P	04/03/2025	14:43	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	14:43	LB135289
CCB04	Aluminum	100	+/-100	U	100	P	04/03/2025	15:36	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	15:36	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	15:36	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	15:36	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	15:36	LB135289
	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	15:36	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	15:36	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	15:36	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	15:36	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	15:36	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	15:36	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	15:36	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1690							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1690	SAS No.: Q1690						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	15:36	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	15:36	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	15:36	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	15:36	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	15:36	LB135289
	Sodium	2000	+/-2000	U	2000	P	04/03/2025	15:36	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	15:36	LB135289
CCB05	Aluminum	100	+/-100	U	100	P	04/03/2025	16:40	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	16:40	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	16:40	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	16:40	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	16:40	LB135289
	Cadmium	0.38	+/-6.00	J	6.00	P	04/03/2025	16:40	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	16:40	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	16:40	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	16:40	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	16:40	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	16:40	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	16:40	LB135289
	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	16:40	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	16:40	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	16:40	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	16:40	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	16:40	LB135289
	Sodium	2000	+/-2000	U	2000	P	04/03/2025	16:40	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	16:40	LB135289
CCB06	Aluminum	100	+/-100	U	100	P	04/03/2025	18:00	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	18:00	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	18:00	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	18:00	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	18:00	LB135289
	Cadmium	6.00	+/-6.00	U	6.00	P	04/03/2025	18:00	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	18:00	LB135289

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1690</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1690</u>	SAS No.: <u>Q1690</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	18:00	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	18:00	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:00	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	18:00	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	18:00	LB135289
	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	18:00	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	18:00	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	18:00	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	18:00	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	18:00	LB135289
	Sodium	624	+/-2000	J	2000	P	04/03/2025	18:00	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:00	LB135289
CCB07	Aluminum	100	+/-100	U	100	P	04/03/2025	18:33	LB135289
	Antimony	50.0	+/-50.0	U	50.0	P	04/03/2025	18:33	LB135289
	Arsenic	20.0	+/-20.0	U	20.0	P	04/03/2025	18:33	LB135289
	Barium	100	+/-100	U	100	P	04/03/2025	18:33	LB135289
	Beryllium	6.00	+/-6.00	U	6.00	P	04/03/2025	18:33	LB135289
	Cadmium	0.23	+/-6.00	J	6.00	P	04/03/2025	18:33	LB135289
	Calcium	2000	+/-2000	U	2000	P	04/03/2025	18:33	LB135289
	Chromium	10.0	+/-10.0	U	10.0	P	04/03/2025	18:33	LB135289
	Cobalt	30.0	+/-30.0	U	30.0	P	04/03/2025	18:33	LB135289
	Copper	20.0	+/-20.0	U	20.0	P	04/03/2025	18:33	LB135289
	Iron	100	+/-100	U	100	P	04/03/2025	18:33	LB135289
	Lead	12.0	+/-12.0	U	12.0	P	04/03/2025	18:33	LB135289
	Magnesium	2000	+/-2000	U	2000	P	04/03/2025	18:33	LB135289
	Manganese	20.0	+/-20.0	U	20.0	P	04/03/2025	18:33	LB135289
	Nickel	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289
	Potassium	2000	+/-2000	U	2000	P	04/03/2025	18:33	LB135289
	Selenium	20.0	+/-20.0	U	20.0	P	04/03/2025	18:33	LB135289
	Silver	10.0	+/-10.0	U	10.0	P	04/03/2025	18:33	LB135289
	Sodium	2000	+/-2000	U	2000	P	04/03/2025	18:33	LB135289
	Thallium	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289
	Vanadium	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289
	Zinc	40.0	+/-40.0	U	40.0	P	04/03/2025	18:33	LB135289