

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q1690
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1690 **SAS No.:** Q1690

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167422BS							
Aluminum	ug/L	1000	931		93	80 - 120	P
Antimony	ug/L	400	403		101	80 - 120	P
Arsenic	ug/L	400	387		97	80 - 120	P
Barium	ug/L	100	87.1		87	80 - 120	P
Beryllium	ug/L	100	92.8		93	80 - 120	P
Cadmium	ug/L	100	93.8		94	80 - 120	P
Calcium	ug/L	500	453	J	91	80 - 120	P
Chromium	ug/L	200	201		100	80 - 120	P
Cobalt	ug/L	100	95.5		96	80 - 120	P
Copper	ug/L	150	155		103	80 - 120	P
Iron	ug/L	1500	1440		96	80 - 120	P
Lead	ug/L	500	467		93	80 - 120	P
Magnesium	ug/L	1000	923	J	92	80 - 120	P
Manganese	ug/L	100	90.4		90	80 - 120	P
Nickel	ug/L	250	241		96	80 - 120	P
Potassium	ug/L	5000	4670		93	80 - 120	P
Selenium	ug/L	1000	983		98	80 - 120	P
Silver	ug/L	37.5	37.4		100	80 - 120	P
Sodium	ug/L	1500	1500		100	80 - 120	P
Thallium	ug/L	1000	1020		102	80 - 120	P
Vanadium	ug/L	150	140		93	80 - 120	P
Zinc	ug/L	100	101		101	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q1690

Contract: GENV01

Lab Code: CHEM

Case No.: Q1690

SAS No.: Q1690

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167425BS Mercury	ug/L	4.0	3.42		86	80 - 120	CV