

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN031825

Review By	John Carlone	Review On	3/19/2025 9:43:20 AM
Supervise By	Maresh Dadoda	Supervise On	3/19/2025 2:22:13 PM
SubDirectory	VN031825	HP Acquire Method	HP Processing Method 82N031825W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133352
Initial Calibration Stds	VP133393,VP133394,VP133395,VP133396,VP133397,VP133399
CCC	VP133353
Internal Standard/PEM	
ICV/I.BLK	VP133401
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085994.D	18 Mar 2025 08:52		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085995.D	18 Mar 2025 10:55	Need ICAL	JC\MD	Not Ok
3	VSTDICC100	VSTDICC100	VN085996.D	18 Mar 2025 11:44		JC\MD	Ok,M
4	VSTDICCC050	VSTDICCC050	VN085997.D	18 Mar 2025 12:08	Comp.#56 and 58 is on Quadratic Regression	JC\MD	Ok,M
5	VSTDICC020	VSTDICC020	VN085998.D	18 Mar 2025 12:32		JC\MD	Ok,M
6	VSTDICC010	VSTDICC010	VN085999.D	18 Mar 2025 12:57	%D failed for Comp. #58 in 01PPB and 20PPB	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN086000.D	18 Mar 2025 13:21		JC\MD	Ok,M
8	VSTDICC001	VSTDICC001	VN086001.D	18 Mar 2025 14:09		JC\MD	Ok,M
9	IBLK	IBLK	VN086002.D	18 Mar 2025 14:34		JC\MD	Ok
10	VSTDICV050	ICVVN031825	VN086003.D	18 Mar 2025 16:49	ICV Failed for comp. #39,52,58,68,69,70,78,80,84,85, 86,89 For DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN040125

Review By	John Carlone	Review On	4/2/2025 9:47:49 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:02 PM
SubDirectory	VN040125	HP Acquire Method	HP Processing Method 82N031825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133553
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133554,VP133558

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086190.D	01 Apr 2025 11:41		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086191.D	01 Apr 2025 12:30	pH#Lot#V12668	JC\MD	Ok,M
3	VN0401MBL01	VN0401MBL01	VN086192.D	01 Apr 2025 13:05		JC\MD	Ok
4	VN0401WBL01	VN0401WBL01	VN086193.D	01 Apr 2025 13:29		JC\MD	Ok
5	VN0401WBS01	VN0401WBS01	VN086194.D	01 Apr 2025 13:54		JC\MD	Ok,M
6	VN0401WBSD01	VN0401WBSD01	VN086195.D	01 Apr 2025 14:48		JC\MD	Ok,M
7	Q1659-01	3808	VN086196.D	01 Apr 2025 15:13	vial A pH<2	JC\MD	Ok
8	Q1659-02	3809	VN086197.D	01 Apr 2025 15:37	RT Shifted ;Internal Standards fail,surrogate fail;Need Straight Run	JC\MD	Not Ok
9	Q1690-01	MW112010	VN086198.D	01 Apr 2025 16:01	vial A pH<2	JC\MD	Ok
10	VSTDCCC050	VSTDCCC050EC	VN086199.D	01 Apr 2025 16:35		JC\MD	Ok,M

M : Manual Integration