

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: **Q1690**

Client: **G Environmental**

Analytical Method: **SW8260D**

Datafile : VN086195.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0401WBSD01	Dichlorodifluoromethane	20	22.5	ug/L	113	8		40 (69)	160 (116)	20 (20)
	Chloromethane	20	21.5	ug/L	108	8		40 (65)	160 (116)	20 (20)
	Vinyl chloride	20	20.7	ug/L	104	6		70 (65)	130 (117)	20 (20)
	Bromomethane	20	20.6	ug/L	103	2		40 (58)	160 (125)	20 (20)
	Chloroethane	20	18.4	ug/L	92	8		40 (56)	160 (128)	20 (20)
	Trichlorofluoromethane	20	17.6	ug/L	88	14		40 (73)	160 (115)	20 (20)
	1,1,2-Trichlorotrifluoroethane	20	17.0	ug/L	85	13		70 (80)	130 (112)	20 (20)
	1,1-Dichloroethene	20	15.2	ug/L	76	18		70 (74)	130 (110)	20 (20)
	Acetone	100	75.7	ug/L	76	8		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	14.6	ug/L	73	14		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	17.6	ug/L	88	3		70 (78)	130 (114)	20 (20)
	Methyl Acetate	20	16.3	ug/L	81	2		70 (67)	130 (125)	20 (20)
	Methylene Chloride	20	17.0	ug/L	85	7		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	16.2	ug/L	81	16		70 (75)	130 (108)	20 (20)
	1,1-Dichloroethane	20	17.0	ug/L	85	8		70 (78)	130 (112)	20 (20)
	Cyclohexane	20	15.8	ug/L	79	8		70 (75)	130 (110)	20 (20)
	2-Butanone	100	81.3	ug/L	81	4		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	19.9	ug/L	100	10		70 (77)	130 (113)	20 (20)
	cis-1,2-Dichloroethene	20	17.4	ug/L	87	6		70 (77)	130 (110)	20 (20)
	Bromochloromethane	20	21.0	ug/L	105	2		70 (70)	130 (124)	20 (20)
	Chloroform	20	19.1	ug/L	96	4		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	19.5	ug/L	98	6		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	17.0	ug/L	85	7		70 (72)	130 (115)	20 (20)
	Benzene	20	18.5	ug/L	93	4		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	20.1	ug/L	101	2		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	17.6	ug/L	88	7		70 (77)	130 (113)	20 (20)
	1,2-Dichloropropane	20	19.1	ug/L	96	0		70 (83)	130 (111)	20 (20)
	Bromodichloromethane	20	20.4	ug/L	102	1		70 (83)	130 (110)	20 (20)
	4-Methyl-2-Pentanone	100	96.0	ug/L	96	2		40 (74)	160 (118)	20 (20)
	Toluene	20	19.7	ug/L	99	7		70 (82)	130 (110)	20 (20)
	t-1,3-Dichloropropene	20	19.4	ug/L	97	0		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	18.7	ug/L	94	4		70 (82)	130 (110)	20 (20)
	1,1,2-Trichloroethane	20	19.8	ug/L	99	5		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	93.5	ug/L	94	6		40 (73)	160 (117)	20 (20)
	Dibromochloromethane	20	20.5	ug/L	103	3		70 (82)	130 (110)	20 (20)
	1,2-Dibromoethane	20	19.7	ug/L	99	1		70 (81)	130 (110)	20 (20)
	Tetrachloroethene	20	19.0	ug/L	95	5		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	17.7	ug/L	89	3		70 (82)	130 (109)	20 (20)
	Ethyl Benzene	20	17.0	ug/L	85	9		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	35.1	ug/L	88	12		70 (82)	130 (110)	20 (20)
	o-Xylene	20	17.8	ug/L	89	11		70 (83)	130 (109)	20 (20)
	Styrene	20	18.0	ug/L	90	9		70 (80)	130 (111)	20 (20)
	Bromoform	20	19.2	ug/L	96	3		70 (79)	130 (109)	20 (20)
	Isopropylbenzene	20	16.5	ug/L	83	7		70 (83)	130 (112)	20 (20)
	1,1,2,2-Tetrachloroethane	20	17.1	ug/L	86	1		70 (76)	130 (118)	20 (20)
	1,3-Dichlorobenzene	20	17.3	ug/L	86	5		70 (82)	130 (108)	20 (20)
	1,4-Dichlorobenzene	20	16.4	ug/L	82	7		70 (82)	130 (107)	20 (20)
	1,2-Dichlorobenzene	20	16.8	ug/L	84	6		70 (82)	130 (109)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1690

Client: G Environmental

Analytical Method: SW8260D

Datafile : VN086195.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0401WBSD01	1,2-Dibromo-3-Chloropropane	20	17.4	ug/L	87	2		40 (68)	160 (112)	20 (20)
	1,2,4-Trichlorobenzene	20	15.0	ug/L	75	3		70 (75)	130 (113)	20 (20)
	1,2,3-Trichlorobenzene	20	15.6	ug/L	78	9		70 (76)	130 (114)	20 (20)