

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1690 SAS No.: Q1690 SDG NO.: Q1690
Lab File ID: VN086191.D Date Analyzed: 04/01/2025
Instrument ID: MSVOA_N Time Analyzed: 12:30
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	131862	8.22	204993	9.10	195897	11.87
UPPER LIMIT	263724	8.724	409986	9.6	391794	12.365
LOWER LIMIT	65931	7.724	102497	8.6	97948.5	11.365
EPA SAMPLE NO.						
MW112010	155880	8.22	278463	9.10	256873	11.87
VN0401WBL01	103379	8.22	183913	9.10	165471	11.87
VN0401WBS01	124765	8.22	197440	9.10	188658	11.87
VN0401WBSD01	113137	8.22	183169	9.10	166660	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.