

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1690 SAS No.: Q1690 SDG NO.: Q1690
 Lab File ID: VN086191.D Date Analyzed: 04/01/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:30
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	111589	13.788				
UPPER LIMIT	223178	14.288				
LOWER LIMIT	55794.5	13.288				
EPA SAMPLE NO.						
MW112010	118313	13.79				
VN0401WBL01	65170	13.79				
VN0401WBS01	101045	13.79				
VN0401WBSD01	87066	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.