

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1690 SAS No.: Q1690 SDG NO.: Q1690
Lab File ID: VN085994.D BFB Injection Date: 03/18/2025
Instrument ID: MSVOA_N BFB Injection Time: 08:52
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.2
75	30.0 - 60.0% of mass 95	52.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.6 (0.7) 1
174	50.0 - 100.0% of mass 95	87.1
175	5.0 - 9.0% of mass 174	6.7 (7.7) 1
176	95.0 - 101.0% of mass 174	83.7 (96.2) 1
177	5.0 - 9.0% of mass 176	5.4 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN085996.D	03/18/2025	11:44
VSTDICCC050	VSTDICCC050	VN085997.D	03/18/2025	12:08
VSTDICC020	VSTDICC020	VN085998.D	03/18/2025	12:32
VSTDICC010	VSTDICC010	VN085999.D	03/18/2025	12:57
VSTDICC005	VSTDICC005	VN086000.D	03/18/2025	13:21
VSTDICC001	VSTDICC001	VN086001.D	03/18/2025	14:09