

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : 416 CLINTON ST HEMPSTEAD

GFE LLC

58 Nokomis Ave

Lake Hiawatha, NJ - 07034

Phone No: 646-542-3465

ORDER ID : Q1691

ATTENTION : Christina Perozzi



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) TO-15- Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) TO-15 Data	8
6) Shipping Document	54
6.1) CHAIN OF CUSTODY	55
6.2) Lab Certificate	59
6.3) Internal COC	60

1
2
3
4
5
6

Cover Page

Order ID : Q1691

Project ID : 416 Clinton St Hempstead

Client : GFE LLC

Lab Sample Number

Q1691-01
Q1691-02
Q1691-03
Q1691-04

Client Sample Number

IA2R
IA3
IA3DUP
OA1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 4/8/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

GFE LLC

Project Name: 416 Clinton St Hempstead

Project # N/A

Chemtech Project # Q1691

Test Name: TO-15

A. Number of Samples and Date of Receipt:

4 Air samples were received on 04/01/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD for {Q1691-01DUP} with File ID: VL042222.D met criteria except for 1,4-Dioxane[27.3%], m/p-Xylene[200%], Tetrachloroethene[20.4%] and Tetrahydrofuran[200%] due to difference in results of original and DUP.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1691

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 04/08/2025

LAB CHRONICLE

OrderID:	Q1691	OrderDate:	4/1/2025 12:37:42 PM
Client:	GFE LLC	Project:	416 Clinton St Hempstead
Contact:	Christina Perozzi	Location:	I41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1691-01	IA2R	Air	TO-15	TO-15	03/31/25		04/04/25	04/01/25
Q1691-02	IA3	Air	TO-15	TO-15	03/31/25		04/04/25	04/01/25
Q1691-03	IA3DUP	Air	TO-15	TO-15	03/31/25		04/04/25	04/01/25
Q1691-04	OA1	Air	TO-15	TO-15	03/31/25		04/04/25	04/01/25

Hit Summary Sheet SW-846

SDG No.: Q1691
Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	IA2R							
Q1691-01	IA2R	Air	Dichlorodifluoromethane	2.03	J	1.04	2.47	ug/m3
Q1691-01	IA2R	Air	Chloromethane	1.92		0.21	1.03	ug/m3
Q1691-01	IA2R	Air	Trichlorofluoromethane	1.18	J	0.90	2.81	ug/m3
Q1691-01	IA2R	Air	Acetone	22.6		0.57	1.19	ug/m3
Q1691-01	IA2R	Air	Methylene Chloride	4.17		0.66	1.74	ug/m3
Q1691-01	IA2R	Air	2-Butanone	3.24		0.35	1.47	ug/m3
Q1691-01	IA2R	Air	Carbon Tetrachloride	0.75		0.060	0.19	ug/m3
Q1691-01	IA2R	Air	Chloroform	4.00		0.39	2.44	ug/m3
Q1691-01	IA2R	Air	Benzene	0.58	J	0.29	1.60	ug/m3
Q1691-01	IA2R	Air	Toluene	1.06	J	0.41	1.88	ug/m3
Q1691-01	IA2R	Air	Tetrachloroethene	1.49		0.14	0.20	ug/m3
Q1691-01	IA2R	Air	Hexane	1.59	J	0.39	1.76	ug/m3
Q1691-01	IA2R	Air	1,4-Dioxane	0.90	J	0.76	1.80	ug/m3
			Total Voc :	45.5				
			Total Concentration:	45.5				
Client ID:	IA3							
Q1691-02	IA3	Air	Dichlorodifluoromethane	1.88	J	1.04	2.47	ug/m3
Q1691-02	IA3	Air	Chloromethane	2.27		0.21	1.03	ug/m3
Q1691-02	IA3	Air	Tetrahydrofuran	0.56	J	0.38	1.47	ug/m3
Q1691-02	IA3	Air	Trichlorofluoromethane	1.24	J	0.90	2.81	ug/m3
Q1691-02	IA3	Air	Acetone	19.7		0.57	1.19	ug/m3
Q1691-02	IA3	Air	Methylene Chloride	1.84		0.66	1.74	ug/m3
Q1691-02	IA3	Air	2-Butanone	1.12	J	0.35	1.47	ug/m3
Q1691-02	IA3	Air	Carbon Tetrachloride	1.32		0.060	0.19	ug/m3
Q1691-02	IA3	Air	Chloroform	3.32		0.39	2.44	ug/m3
Q1691-02	IA3	Air	2,2,4-Trimethylpentane	0.47	J	0.47	2.34	ug/m3
Q1691-02	IA3	Air	Benzene	0.73	J	0.29	1.60	ug/m3
Q1691-02	IA3	Air	Toluene	1.21	J	0.41	1.88	ug/m3
Q1691-02	IA3	Air	Tetrachloroethene	0.54		0.14	0.20	ug/m3
Q1691-02	IA3	Air	m/p-Xylene	0.91	J	0.91	4.34	ug/m3
Q1691-02	IA3	Air	Styrene	0.64	J	0.47	2.13	ug/m3
Q1691-02	IA3	Air	Hexane	1.20	J	0.39	1.76	ug/m3
			Total Voc :	39.0				
			Total Concentration:	39.0				
Client ID:	IA3DUP							
Q1691-03	IA3DUP	Air	Dichlorodifluoromethane	2.03	J	1.04	2.47	ug/m3

Hit Summary Sheet SW-846

SDG No.: Q1691

Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q1691-03	IA3DUP	Air	Chloromethane	1.71		0.21	1.03	ug/m3
Q1691-03	IA3DUP	Air	Tetrahydrofuran	0.47	J	0.38	1.47	ug/m3
Q1691-03	IA3DUP	Air	Trichlorofluoromethane	1.18	J	0.90	2.81	ug/m3
Q1691-03	IA3DUP	Air	Acetone	15.2		0.57	1.19	ug/m3
Q1691-03	IA3DUP	Air	Methylene Chloride	2.95		0.66	1.74	ug/m3
Q1691-03	IA3DUP	Air	2-Butanone	1.03	J	0.35	1.47	ug/m3
Q1691-03	IA3DUP	Air	Carbon Tetrachloride	1.13		0.060	0.19	ug/m3
Q1691-03	IA3DUP	Air	Chloroform	3.61		0.39	2.44	ug/m3
Q1691-03	IA3DUP	Air	2,2,4-Trimethylpentane	0.47	J	0.47	2.34	ug/m3
Q1691-03	IA3DUP	Air	Benzene	0.61	J	0.29	1.60	ug/m3
Q1691-03	IA3DUP	Air	Toluene	1.28	J	0.41	1.88	ug/m3
Q1691-03	IA3DUP	Air	Tetrachloroethene	2.03		0.14	0.20	ug/m3
Q1691-03	IA3DUP	Air	m/p-Xylene	0.96	J	0.91	4.34	ug/m3
Q1691-03	IA3DUP	Air	Styrene	0.60	J	0.47	2.13	ug/m3
Q1691-03	IA3DUP	Air	Hexane	1.90		0.39	1.76	ug/m3
			Total Voc :		37.2			
			Total Concentration:		37.2			
Client ID:	OA1							
Q1691-04	OA1	Air	Dichlorodifluoromethane	2.03	J	1.04	2.47	ug/m3
Q1691-04	OA1	Air	Chloromethane	1.12		0.21	1.03	ug/m3
Q1691-04	OA1	Air	Trichlorofluoromethane	1.18	J	0.90	2.81	ug/m3
Q1691-04	OA1	Air	Acetone	10.7		0.57	1.19	ug/m3
Q1691-04	OA1	Air	Methylene Chloride	1.70	J	0.66	1.74	ug/m3
Q1691-04	OA1	Air	2-Butanone	1.00	J	0.35	1.47	ug/m3
Q1691-04	OA1	Air	Carbon Tetrachloride	0.44		0.060	0.19	ug/m3
Q1691-04	OA1	Air	Benzene	0.51	J	0.29	1.60	ug/m3
Q1691-04	OA1	Air	Toluene	1.17	J	0.41	1.88	ug/m3
Q1691-04	OA1	Air	Tetrachloroethene	2.51		0.14	0.20	ug/m3
Q1691-04	OA1	Air	Hexane	0.81	J	0.39	1.76	ug/m3
			Total Voc :		23.2			
			Total Concentration:		23.2			

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : IA2R

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-01

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.41	J	2.03		
Chloromethane	74-87-3	50.49	0.93		1.92		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	0.13	U	0.38		
Trichlorofluoromethane	75-69-4	137.4	0.21	J	1.18		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.19	U	0.58		
Heptane	142-82-5	100.2	0.11	U	0.45		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	9.5		22.6		
Carbon Disulfide	75-15-0	76.14	0.16	U	0.5		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	1.2		4.17		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	1.1		3.24		
Carbon Tetrachloride	56-23-5	153.8	0.12		0.75		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.82		4		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.1	U	0.47		
Benzene	71-43-2	78.11	0.18	J	0.58		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.09	U	0.37		
Toluene	108-88-3	92.14	0.28	J	1.06		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : IA2R

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-01

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	0.22		1.49		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.21	U	0.91		
o-Xylene	95-47-6	106.2	0.12	U	0.52		
Styrene	100-42-5	104.1	0.11	U	0.47		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.08	U	0.39		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	0.05	U	0.3		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	0.45	J	1.59		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.25	J	0.9		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : IA3

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-02

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.38	J	1.88		
Chloromethane	74-87-3	50.49	1.1		2.27		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	0.19	J	0.56		
Trichlorofluoromethane	75-69-4	137.4	0.22	J	1.24		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.19	U	0.58		
Heptane	142-82-5	100.2	0.11	U	0.45		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	8.3		19.7		
Carbon Disulfide	75-15-0	76.14	0.16	U	0.5		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	0.53		1.84		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.38	J	1.12		
Carbon Tetrachloride	56-23-5	153.8	0.21		1.32		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.68		3.32		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.1	J	0.47		
Benzene	71-43-2	78.11	0.23	J	0.73		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.09	U	0.37		
Toluene	108-88-3	92.14	0.32	J	1.21		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : IA3

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-02

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	0.08		0.54		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.21	J	0.91		
o-Xylene	95-47-6	106.2	0.12	U	0.52		
Styrene	100-42-5	104.1	0.15	J	0.64		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.08	U	0.39		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	0.05	U	0.3		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	0.34	J	1.2		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : IA3DUP

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-03

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.41	J	2.03		
Chloromethane	74-87-3	50.49	0.83		1.71		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	0.16	J	0.47		
Trichlorofluoromethane	75-69-4	137.4	0.21	J	1.18		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.19	U	0.58		
Heptane	142-82-5	100.2	0.11	U	0.45		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	6.4		15.2		
Carbon Disulfide	75-15-0	76.14	0.16	U	0.5		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	0.85		2.95		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.35	J	1.03		
Carbon Tetrachloride	56-23-5	153.8	0.18		1.13		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.74		3.61		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.1	J	0.47		
Benzene	71-43-2	78.11	0.19	J	0.61		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.09	U	0.37		
Toluene	108-88-3	92.14	0.34	J	1.28		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : IA3DUP

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-03

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	0.3		2.03		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.22	J	0.96		
o-Xylene	95-47-6	106.2	0.12	U	0.52		
Styrene	100-42-5	104.1	0.14	J	0.6		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.08	U	0.39		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	0.05	U	0.3		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	0.54		1.9		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : OA1

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-04

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Dichlorodifluoromethane	75-71-8	120.9	0.41	J	2.03		
Chloromethane	74-87-3	50.49	0.54		1.12		
Vinyl Chloride	75-01-4	62.5	0.01	U	0.03		
Bromomethane	74-83-9	94.94	0.14	U	0.54		
Chloroethane	75-00-3	64.52	0.17	U	0.45		
Tetrahydrofuran	109-99-9	72.11	0.13	U	0.38		
Trichlorofluoromethane	75-69-4	137.4	0.21	J	1.18		
Dichlorotetrafluoroethane	76-14-2	170.9	0.09	U	0.63		
1,1,2-Trichlorotrifluoroethane	76-13-1	187.4	0.16	U	1.23		
tert-Butyl alcohol	75-65-0	74.12	0.19	U	0.58		
Heptane	142-82-5	100.2	0.11	U	0.45		
1,1-Dichloroethene	75-35-4	96.94	0.14	U	0.56		
Acetone	67-64-1	58.08	4.5		10.7		
Carbon Disulfide	75-15-0	76.14	0.16	U	0.5		
Methyl tert-Butyl Ether	1634-04-4	88.15	0.12	U	0.43		
Methylene Chloride	75-09-2	84.94	0.49	J	1.7		
trans-1,2-Dichloroethene	156-60-5	96.94	0.19	U	0.75		
1,1-Dichloroethane	75-34-3	98.96	0.13	U	0.53		
Cyclohexane	110-82-7	84.16	0.22	U	0.76		
2-Butanone	78-93-3	72.11	0.34	J	1		
Carbon Tetrachloride	56-23-5	153.8	0.07		0.44		
cis-1,2-Dichloroethene	156-59-2	96.94	0.09	U	0.36		
Chloroform	67-66-3	119.4	0.08	U	0.39		
1,1,1-Trichloroethane	71-55-6	133.4	0.01	U	0.05		
2,2,4-Trimethylpentane	540-84-1	114.2	0.1	U	0.47		
Benzene	71-43-2	78.11	0.16	J	0.51		
1,2-Dichloroethane	107-06-2	98.96	0.09	U	0.36		
Trichloroethene	79-01-6	131.4	0.02	U	0.11		
1,2-Dichloropropane	78-87-5	113	0.11	U	0.51		
Bromodichloromethane	75-27-4	163.8	0.08	U	0.54		
4-Methyl-2-Pentanone	108-10-1	100.2	0.09	U	0.37		
Toluene	108-88-3	92.14	0.31	J	1.17		
t-1,3-Dichloropropene	10061-02-6	111	0.06	U	0.27		
cis-1,3-Dichloropropene	10061-01-5	111	0.06	U	0.27		
1,1,2-Trichloroethane	79-00-5	133.4	0.07	U	0.38		

Project : 416 Clinton St Hempstead

Sampling Date : 03/31/25

Field Id Number : OA1

Analysis Date : 04/04/25

Laboratory Id Number : Q1691-04

Target Analyts : Air Results

Dibromochloromethane	124-48-1	208.3	0.07	U	0.6		
1,2-Dibromoethane	106-93-4	187.9	0.07	U	0.54		
Tetrachloroethene	127-18-4	165.8	0.37		2.51		
Chlorobenzene	108-90-7	112.6	0.08	U	0.37		
Ethyl Benzene	100-41-4	106.2	0.12	U	0.52		
m/p-Xylene	179601-23-1	106.2	0.21	U	0.91		
o-Xylene	95-47-6	106.2	0.12	U	0.52		
Styrene	100-42-5	104.1	0.11	U	0.47		
Bromoform	75-25-2	252.8	0.06	U	0.62		
1,1,2,2-Tetrachloroethane	79-34-5	167.9	0.01	U	0.07		
2-Chlorotoluene	95-49-8	126.6	0.11	U	0.57		
1,3,5-Trimethylbenzene	108-67-8	120.2	0.11	U	0.54		
1,2,4-Trimethylbenzene	95-63-6	120.2	0.08	U	0.39		
1,3-Dichlorobenzene	541-73-1	147	0.08	U	0.48		
1,4-Dichlorobenzene	106-46-7	147	0.05	U	0.3		
1,2-Dichlorobenzene	95-50-1	147	0.08	U	0.48		
1,2,4-Trichlorobenzene	120-82-1	181.5	0.09	U	0.67		
Hexachloro-1,3-Butadiene	87-68-3	260.8	0.09	U	0.96		
Naphthalene	91-20-3	128.17	0.08	U	0.42		
1,3-Butadiene	106-99-0	54.09	0.13	U	0.29		
4-Ethyltoluene	622-96-8	120.2	0.12	U	0.59		
Hexane	110-54-3	86.17	0.23	J	0.81		
Allyl Chloride	107-05-1	76.53	0.15	U	0.47		
1,4-Dioxane	123-91-1	88.12	0.21	U	0.76		
Methyl Methacrylate	80-62-6	100.117	0.09	U	0.37		



SAMPLE DATA

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA2R	SDG No.:	Q1691
Lab Sample ID:	Q1691-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042221.D	1		04/04/25 12:46	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.41	2.03	J	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.93	1.92		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.21	1.18	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	9.50	22.6		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	1.20	4.17		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	1.10	3.24		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.12	0.75		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.82	4.00		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.18	0.58	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.28	1.06	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA2R	SDG No.:	Q1691
Lab Sample ID:	Q1691-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042221.D	1		04/04/25 12:46	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.22	1.49		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.45	1.59	J	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.25	0.90	J	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.80			65 - 135	98%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	194000		2.787			
540-36-3	1,4-Difluorobenzene	537000		3.959			
3114-55-4	Chlorobenzene-d5	466000		8.882			

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA2R	SDG No.:	Q1691
Lab Sample ID:	Q1691-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042221.D	1		04/04/25 12:46	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA3	SDG No.:	Q1691
Lab Sample ID:	Q1691-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042224.D	1		04/04/25 14:26	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.38	1.88	J	1.04	2.47	ug/m3
74-87-3	Chloromethane	1.10	2.27		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.19	0.56	J	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.22	1.24	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	8.30	19.7		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.53	1.84		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.38	1.12	J	0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.21	1.32		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.68	3.32		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	J	0.47	2.34	ug/m3
71-43-2	Benzene	0.23	0.73	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.32	1.21	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA3	SDG No.:	Q1691
Lab Sample ID:	Q1691-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042224.D	1		04/04/25 14:26	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.080	0.54		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.15	0.64	J	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.34	1.20	J	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.80			65 - 135	98%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	191000		2.787			
540-36-3	1,4-Difluorobenzene	505000		3.959			
3114-55-4	Chlorobenzene-d5	449000		8.882			

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA3	SDG No.:	Q1691
Lab Sample ID:	Q1691-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042224.D	1		04/04/25 14:26	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA3DUP	SDG No.:	Q1691
Lab Sample ID:	Q1691-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042225.D	1		04/04/25 15:50	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.41	2.03	J	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.83	1.71		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.16	0.47	J	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.21	1.18	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	6.40	15.2		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.85	2.95		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.35	1.03	J	0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.18	1.13		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.74	3.61		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	J	0.47	2.34	ug/m3
71-43-2	Benzene	0.19	0.61	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.34	1.28	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA3DUP	SDG No.:	Q1691
Lab Sample ID:	Q1691-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042225.D	1		04/04/25 15:50	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.30	2.03		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.22	0.96	J	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.14	0.60	J	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.54	1.90		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.80			65 - 135	98%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	192000		2.787			
540-36-3	1,4-Difluorobenzene	513000		3.955			
3114-55-4	Chlorobenzene-d5	454000		8.882			

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	IA3DUP	SDG No.:	Q1691
Lab Sample ID:	Q1691-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042225.D	1		04/04/25 15:50	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	OA1	SDG No.:	Q1691
Lab Sample ID:	Q1691-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042226.D	1		04/04/25 16:24	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.41	2.03	J	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.54	1.12		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.21	1.18	J	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	4.50	10.7		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.49	1.70	J	0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.34	1.00	J	0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.070	0.44		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.080	0.39	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.16	0.51	J	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.31	1.17	J	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	OA1	SDG No.:	Q1691
Lab Sample ID:	Q1691-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042226.D	1		04/04/25 16:24	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.37	2.51		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.23	0.81	J	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.80			65 - 135	98%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	195000		2.784			
540-36-3	1,4-Difluorobenzene	518000		3.955			
3114-55-4	Chlorobenzene-d5	453000		8.878			

Report of Analysis

Client:	GFE LLC	Date Collected:	03/31/25
Project:	416 Clinton St Hempstead	Date Received:	04/01/25
Client Sample ID:	OA1	SDG No.:	Q1691
Lab Sample ID:	Q1691-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042226.D	1		04/04/25 16:24	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements



QC SUMMARY

Surrogate Summary

SDG No.: Q1691

Client: GFE LLC

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1691-01	IA2R	1-Bromo-4-Fluorobenzene	10	9.83	98	65	135
Q1691-01DUP	IA2RDUP	1-Bromo-4-Fluorobenzene	10	9.90	99	65	135
Q1691-02	IA3	1-Bromo-4-Fluorobenzene	10	9.78	98	65	135
Q1691-03	IA3DUP	1-Bromo-4-Fluorobenzene	10	9.82	98	65	135
Q1691-04	OA1	1-Bromo-4-Fluorobenzene	10	9.80	98	65	135
VL0404ABL01	VL0404ABL01	1-Bromo-4-Fluorobenzene	10	9.56	96	65	135
VL0404ABS01	VL0404ABS01	1-Bromo-4-Fluorobenzene	10	10.2	102	65	135

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1691

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042220.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0404ABS01	Dichlorodifluoromethane	10	9.80	ppbv	98			70	130	
	Chloromethane	10	9.00	ppbv	90			70	130	
	Vinyl Chloride	10	8.20	ppbv	82			70	130	
	Bromomethane	10	8.50	ppbv	85			70	130	
	Chloroethane	10	8.70	ppbv	87			70	130	
	Tetrahydrofuran	10	8.80	ppbv	88			70	130	
	Trichlorofluoromethane	10	9.70	ppbv	97			70	130	
	1,1,2-Trichlorotrifluoroethane	10	8.90	ppbv	89			70	130	
	Dichlorotetrafluoroethane	10	9.70	ppbv	97			70	130	
	tert-Butyl Alcohol	10	8.30	ppbv	83			70	130	
	Heptane	10	8.50	ppbv	85			70	130	
	1,1-Dichloroethene	10	8.80	ppbv	88			70	130	
	Acetone	10	8.40	ppbv	84			70	130	
	Carbon disulfide	10	8.50	ppbv	85			70	130	
	Methyl tert-butyl Ether	10	8.50	ppbv	85			70	130	
	Methylene Chloride	10	7.90	ppbv	79			70	130	
	trans-1,2-Dichloroethene	10	8.70	ppbv	87			70	130	
	1,1-Dichloroethane	10	9.40	ppbv	94			70	130	
	Cyclohexane	10	8.50	ppbv	85			70	130	
	2-Butanone	10	9.20	ppbv	92			70	130	
	Carbon Tetrachloride	10	10.3	ppbv	103			70	130	
	cis-1,2-Dichloroethene	10	9.00	ppbv	90			70	130	
	Chloroform	10	10.1	ppbv	101			70	130	
	1,1,1-Trichloroethane	10	9.10	ppbv	91			70	130	
	2,2,4-Trimethylpentane	10	9.30	ppbv	93			70	130	
	Benzene	10	9.20	ppbv	92			70	130	
	1,2-Dichloroethane	10	10.6	ppbv	106			70	130	
	Trichloroethene	10	9.10	ppbv	91			70	130	
	1,2-Dichloropropane	10	9.10	ppbv	91			70	130	
	Bromodichloromethane	10	10.5	ppbv	105			70	130	
	4-Methyl-2-Pentanone	10	8.40	ppbv	84			70	130	
	Toluene	10	9.00	ppbv	90			70	130	
	t-1,3-Dichloropropene	10	8.20	ppbv	82			70	130	
	cis-1,3-Dichloropropene	10	8.40	ppbv	84			70	130	
	1,1,2-Trichloroethane	10	9.90	ppbv	99			70	130	
	Dibromochloromethane	10	10.2	ppbv	102			70	130	
	1,2-Dibromoethane	10	9.80	ppbv	98			70	130	
	Tetrachloroethene	10	8.50	ppbv	85			70	130	
	Chlorobenzene	10	10.1	ppbv	101			70	130	
	Ethyl Benzene	10	9.80	ppbv	98			70	130	
	m/p-Xylene	20	20.4	ppbv	102			70	130	
	o-Xylene	10	10.2	ppbv	102			70	130	
	Styrene	10	9.50	ppbv	95			70	130	
	Bromoform	10	9.90	ppbv	99			70	130	
	1,1,2,2-Tetrachloroethane	10	10.7	ppbv	107			70	130	
	2-Chlorotoluene	10	10.3	ppbv	103			70	130	
	1,3,5-Trimethylbenzene	10	10.4	ppbv	104			70	130	
	1,2,4-Trimethylbenzene	10	10.0	ppbv	100			70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1691

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042220.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0404ABS01	1,3-Dichlorobenzene	10	10.7	ppbv	107			70	130	
	1,4-Dichlorobenzene	10	10.7	ppbv	107			70	130	
	1,2-Dichlorobenzene	10	10.8	ppbv	108			70	130	
	1,2,4-Trichlorobenzene	10	9.20	ppbv	92			70	130	
	Hexachloro-1,3-butadiene	10	8.30	ppbv	83			70	130	
	Naphthalene	10	9.50	ppbv	95			70	130	
	1,3-Butadiene	10	8.60	ppbv	86			70	130	
	4-Ethyltoluene	10	10.5	ppbv	105			70	130	
	Hexane	10	8.80	ppbv	88			70	130	
	Allyl Chloride	10	8.60	ppbv	86			70	130	
	1,4-Dioxane	10	8.70	ppbv	87			70	130	
	Methyl methacrylate	10	8.60	ppbv	86			70	130	

Duplicate Sample Summary

Lab Sample Id :	Q1691-01DUP	Q1691-01
Client Id :	IA2RDUP	IA2R
DF :	1	1
Datafile :	VL042222.D	VL042221.D
Anal Date & Time :	04/04/2025 13:20	04/04/2025 12:46

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0.19	0.25	27.3 *
2,2,4-Trimethylpentane	0	0	0
2-Butanone	1.1	1.1	0
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	9.7	9.5	2.1
Allyl Chloride	0	0	0
Benzene	0.18	0.18	0
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.11	0.12	8.7

Duplicate Sample Summary

Lab Sample Id :	Q1691-01DUP	Q1691-01
Client Id :	IA2RDUP	IA2R
DF :	1	1
Datafile :	VL042222.D	VL042221.D
Anal Date & Time :	04/04/2025 13:20	04/04/2025 12:46

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0.84	0.82	2.4
Chloromethane	0.96	0.93	3.2
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.41	0.41	0
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0.42	0.45	6.9
m/p-Xylene	0.17	0	200 *
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	1	1.2	18.2
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0.27	0.22	20.4 *
Tetrahydrofuran	0.1	0	200 *
Toluene	0.28	0.28	0
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.23	0.21	9.1
Vinyl Chloride	0	0	0

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VL0404ABL01

Lab Name: CHEMTECH

Contract: GFEL01

Lab Code: CHEM Case No.: Q1691

SAS No.: Q1691 SDG NO.: Q1691

Lab File ID: VL042216.D

Lab Sample ID: VL0404ABL01

Date Analyzed: 04/04/2025

Time Analyzed: 09:41

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_L

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VL0404ABS01	VL0404ABS01	VL042220.D	04/04/2025
IA2R	Q1691-01	VL042221.D	04/04/2025
IA2RDUP	Q1691-01DUP	VL042222.D	04/04/2025
IA3	Q1691-02	VL042224.D	04/04/2025
IA3DUP	Q1691-03	VL042225.D	04/04/2025
OA1	Q1691-04	VL042226.D	04/04/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GFEL01
Lab Code: CHEM Case No.: Q1691 SAS No.: Q1691 SDG NO.: Q1691
Lab File ID: VL042171.D BFB Injection Date: 03/27/2025
Instrument ID: MSVOA_L BFB Injection Time: 07:44
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	23
75	30.0 - 66.0% of mass 95	56.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 120.0% of mass 95	63.8
175	4.0 - 9.0% of mass 174	5.1 (8) 1
176	93.0 - 101.0% of mass 174	61.1 (95.7) 1
177	5.0 - 9.0% of mass 176	4.2 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICCC010	VSTDICCC010	VL042172.D	03/27/2025	08:26
VSTDICCC002	VSTDICCC002	VL042173.D	03/27/2025	08:59
VSTDICCC001	VSTDICCC001	VL042174.D	03/27/2025	09:30
VSTDICCC0.5	VSTDICCC0.5	VL042175.D	03/27/2025	10:01
VSTDICCC0.1	VSTDICCC0.1	VL042176.D	03/27/2025	10:32
VSTDICCC0.03	VSTDICCC0.03	VL042177.D	03/27/2025	11:03
VSTDICCC015	VSTDICCC015	VL042178.D	03/27/2025	11:36

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GFEL01
Lab Code: CHEM Case No.: Q1691 SAS No.: Q1691 SDG NO.: Q1691
Lab File ID: VL042214.D BFB Injection Date: 04/04/2025
Instrument ID: MSVOA_L BFB Injection Time: 08:05
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	25.3
75	30.0 - 66.0% of mass 95	57.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.3 (0.4) 1
174	50.0 - 120.0% of mass 95	61.2
175	4.0 - 9.0% of mass 174	4.5 (7.3) 1
176	93.0 - 101.0% of mass 174	57.3 (93.5) 1
177	5.0 - 9.0% of mass 176	3.6 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL042215.D	04/04/2025	08:58
VL0404ABL01	VL0404ABL01	VL042216.D	04/04/2025	09:41
VL0404ABS01	VL0404ABS01	VL042220.D	04/04/2025	11:55
IA2R	Q1691-01	VL042221.D	04/04/2025	12:46
IA2RDUP	Q1691-01DUP	VL042222.D	04/04/2025	13:20
IA3	Q1691-02	VL042224.D	04/04/2025	14:26
IA3DUP	Q1691-03	VL042225.D	04/04/2025	15:50
OA1	Q1691-04	VL042226.D	04/04/2025	16:24

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01
Lab Code: CHEM Case No.: Q1691 SAS No.: Q1691 SDG NO.: Q1691
Lab File ID: VL042215.D Date Analyzed: 04/04/2025
Instrument ID: MSVOA_L Time Analyzed: 08:58
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	187369	2.79	519347	3.96	464975	8.89
UPPER LIMIT	262317	3.12	727086	4.29	650965	9.22
LOWER LIMIT	112421	2.46	311608	3.63	278985	8.56
EPA SAMPLE NO.						
IA2R	193944	2.79	537035	3.96	465935	8.88
IA2RDUP	190677	2.79	525113	3.96	459228	8.89
IA3	191073	2.79	505458	3.96	449107	8.88
IA3DUP	191637	2.79	513363	3.96	453956	8.88
OA1	194580	2.78	518092	3.96	453107	8.88
VL0404ABL01	191004	2.79	523611	3.96	464191	8.89
VL0404ABS01	185704	2.79	509064	3.96	456962	8.89

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	416 Clinton St Hempstead	Date Received:	
Client Sample ID:	VL0404ABL01	SDG No.:	Q1691
Lab Sample ID:	VL0404ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042216.D	1		04/04/25 09:41	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.21	1.04	U	1.04	2.47	ug/m3
74-87-3	Chloromethane	0.10	0.21	U	0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	0.010	0.030	U	0.030	0.080	ug/m3
74-83-9	Bromomethane	0.14	0.54	U	0.54	1.94	ug/m3
75-00-3	Chloroethane	0.17	0.45	U	0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	0.13	0.38	U	0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	0.16	0.90	U	0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.16	1.23	U	1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.090	0.63	U	0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	0.19	0.58	U	0.58	1.52	ug/m3
142-82-5	Heptane	0.11	0.45	U	0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	0.14	0.56	U	0.56	1.98	ug/m3
67-64-1	Acetone	0.24	0.57	U	0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	0.16	0.50	U	0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.12	0.43	U	0.43	1.80	ug/m3
75-09-2	Methylene Chloride	0.19	0.66	U	0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.19	0.75	U	0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02	ug/m3
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72	ug/m3
78-93-3	2-Butanone	0.12	0.35	U	0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	0.010	0.060	U	0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.090	0.36	U	0.36	1.98	ug/m3
67-66-3	Chloroform	0.080	0.39	U	0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	0.010	0.050	U	0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.10	0.47	U	0.47	2.34	ug/m3
71-43-2	Benzene	0.090	0.29	U	0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02	ug/m3
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	0.11	0.51	U	0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	0.080	0.54	U	0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.090	0.37	U	0.37	2.05	ug/m3
108-88-3	Toluene	0.11	0.41	U	0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.060	0.27	U	0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	0.070	0.38	U	0.38	2.73	ug/m3

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	416 Clinton St Hempstead	Date Received:	
Client Sample ID:	VL0404ABL01	SDG No.:	Q1691
Lab Sample ID:	VL0404ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042216.D	1		04/04/25 09:41	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.070	0.60	U	0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	0.070	0.54	U	0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20	ug/m3
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	0.12	0.52	U	0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	0.21	0.91	U	0.91	4.34	ug/m3
95-47-6	o-Xylene	0.12	0.52	U	0.52	2.17	ug/m3
100-42-5	Styrene	0.11	0.47	U	0.47	2.13	ug/m3
75-25-2	Bromoform	0.060	0.62	U	0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.010	0.070	U	0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	0.11	0.57	U	0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.11	0.54	U	0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.080	0.39	U	0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	0.050	0.30	U	0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	0.080	0.48	U	0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	0.090	0.67	U	0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.090	0.96	U	0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	0.13	0.29	U	0.29	1.11	ug/m3
91-20-3	Naphthalene	0.080	0.42	U	0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	0.12	0.59	U	0.59	2.46	ug/m3
110-54-3	Hexane	0.11	0.39	U	0.39	1.76	ug/m3
107-05-1	Allyl Chloride	0.15	0.47	U	0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	0.21	0.76	U	0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	0.090	0.37	U	0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	9.60			65 - 135	96%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	191000		2.787			
540-36-3	1,4-Difluorobenzene	524000		3.962			
3114-55-4	Chlorobenzene-d5	464000		8.885			

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	416 Clinton St Hempstead	Date Received:	
Client Sample ID:	VL0404ABL01	SDG No.:	Q1691
Lab Sample ID:	VL0404ABL01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042216.D	1		04/04/25 09:41	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	416 Clinton St Hempstead	Date Received:	
Client Sample ID:	VL0404ABS01	SDG No.:	Q1691
Lab Sample ID:	VL0404ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042220.D	1		04/04/25 11:55	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	9.80	48.5		1.04	2.47	ug/m3
74-87-3	Chloromethane	9.00	18.6		0.21	1.03	ug/m3
75-01-4	Vinyl Chloride	8.20	21.0		0.030	0.080	ug/m3
74-83-9	Bromomethane	8.50	33.0		0.54	1.94	ug/m3
75-00-3	Chloroethane	8.70	23.0		0.45	1.32	ug/m3
109-99-9	Tetrahydrofuran	8.80	25.9		0.38	1.47	ug/m3
75-69-4	Trichlorofluoromethane	9.70	54.5		0.90	2.81	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	8.90	68.2		1.23	3.83	ug/m3
76-14-2	Dichlorotetrafluoroethane	9.70	67.8		0.63	3.49	ug/m3
75-65-0	tert-Butyl alcohol	8.30	25.2		0.58	1.52	ug/m3
142-82-5	Heptane	8.50	34.8		0.45	2.05	ug/m3
75-35-4	1,1-Dichloroethene	8.80	34.9		0.56	1.98	ug/m3
67-64-1	Acetone	8.40	19.9		0.57	1.19	ug/m3
75-15-0	Carbon Disulfide	8.50	26.5		0.50	1.56	ug/m3
1634-04-4	Methyl tert-Butyl Ether	8.50	30.6		0.43	1.80	ug/m3
75-09-2	Methylene Chloride	7.90	27.4		0.66	1.74	ug/m3
156-60-5	trans-1,2-Dichloroethene	8.70	34.5		0.75	1.98	ug/m3
75-34-3	1,1-Dichloroethane	9.40	38.0		0.53	2.02	ug/m3
110-82-7	Cyclohexane	8.50	29.3		0.76	1.72	ug/m3
78-93-3	2-Butanone	9.20	27.1		0.35	1.47	ug/m3
56-23-5	Carbon Tetrachloride	10.3	64.8		0.060	0.19	ug/m3
156-59-2	cis-1,2-Dichloroethene	9.00	35.7		0.36	1.98	ug/m3
67-66-3	Chloroform	10.1	49.3		0.39	2.44	ug/m3
71-55-6	1,1,1-Trichloroethane	9.10	49.6		0.050	0.16	ug/m3
540-84-1	2,2,4-Trimethylpentane	9.30	43.4		0.47	2.34	ug/m3
71-43-2	Benzene	9.20	29.4		0.29	1.60	ug/m3
107-06-2	1,2-Dichloroethane	10.6	42.9		0.36	2.02	ug/m3
79-01-6	Trichloroethene	9.10	48.9		0.11	0.16	ug/m3
78-87-5	1,2-Dichloropropane	9.10	42.1		0.51	2.31	ug/m3
75-27-4	Bromodichloromethane	10.5	70.3		0.54	3.35	ug/m3
108-10-1	4-Methyl-2-Pentanone	8.40	34.4		0.37	2.05	ug/m3
108-88-3	Toluene	9.00	33.9		0.41	1.88	ug/m3
10061-02-6	t-1,3-Dichloropropene	8.20	37.2		0.27	2.27	ug/m3
10061-01-5	cis-1,3-Dichloropropene	8.40	38.1		0.27	2.27	ug/m3
79-00-5	1,1,2-Trichloroethane	9.90	54.0		0.38	2.73	ug/m3

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	416 Clinton St Hempstead	Date Received:	
Client Sample ID:	VL0404ABS01	SDG No.:	Q1691
Lab Sample ID:	VL0404ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042220.D	1		04/04/25 11:55	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	10.2	86.9		0.60	4.26	ug/m3
106-93-4	1,2-Dibromoethane	9.80	75.3		0.54	0.77	ug/m3
127-18-4	Tetrachloroethene	8.50	57.6		0.14	0.20	ug/m3
108-90-7	Chlorobenzene	10.1	46.5		0.37	2.30	ug/m3
100-41-4	Ethyl Benzene	9.80	42.6		0.52	2.17	ug/m3
179601-23-1	m/p-Xylene	20.4	88.6		0.91	4.34	ug/m3
95-47-6	o-Xylene	10.2	44.3		0.52	2.17	ug/m3
100-42-5	Styrene	9.50	40.5		0.47	2.13	ug/m3
75-25-2	Bromoform	9.90	102		0.62	5.17	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	10.7	73.5		0.070	0.21	ug/m3
95-49-8	2-Chlorotoluene	10.3	53.3		0.57	2.59	ug/m3
108-67-8	1,3,5-Trimethylbenzene	10.4	51.1		0.54	2.46	ug/m3
95-63-6	1,2,4-Trimethylbenzene	10.0	49.2		0.39	2.46	ug/m3
541-73-1	1,3-Dichlorobenzene	10.7	64.3		0.48	3.01	ug/m3
106-46-7	1,4-Dichlorobenzene	10.7	64.3		0.30	3.01	ug/m3
95-50-1	1,2-Dichlorobenzene	10.8	64.9		0.48	3.01	ug/m3
120-82-1	1,2,4-Trichlorobenzene	9.20	68.3		0.67	3.71	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	8.30	88.5		0.96	5.33	ug/m3
106-99-0	1,3-Butadiene	8.60	19.0		0.29	1.11	ug/m3
91-20-3	Naphthalene	9.50	49.8		0.42	0.52	ug/m3
622-96-8	4-Ethyltoluene	10.5	51.6		0.59	2.46	ug/m3
110-54-3	Hexane	8.80	31.0		0.39	1.76	ug/m3
107-05-1	Allyl Chloride	8.60	26.9		0.47	1.57	ug/m3
123-91-1	1,4-Dioxane	8.70	31.4		0.76	1.80	ug/m3
80-62-6	Methyl Methacrylate	8.60	35.2		0.37	2.05	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.2			65 - 135	102%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	186000			2.79		
540-36-3	1,4-Difluorobenzene	509000			3.962		
3114-55-4	Chlorobenzene-d5	457000			8.888		

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	416 Clinton St Hempstead	Date Received:	
Client Sample ID:	VL0404ABS01	SDG No.:	Q1691
Lab Sample ID:	VL0404ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042220.D	1		04/04/25 11:55	VL040425

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	---------------	----------------	-----------	-----	------------	-------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements



CALIBRATION SUMMARY

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL032725AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Fri Mar 28 02:00:52 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL042177.D 0.1 =VL042176.D 0.5 =VL042175.D 1 =VL042174.D 2 =VL042173.D 10 =VL042172.D 15 =VL042178.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.602	1.514	1.482	1.404	1.380	1.476	6.03
3) Chlorodifluoro...			1.622	1.541	1.514	1.457	1.408	1.508	5.41
4) Chloromethane			0.660	0.597	0.618	0.573	0.570	0.603	6.17
5) T Vinyl Chloride	0.882	0.705	0.635	0.576	0.584	0.549	0.546	0.639	18.84
6) T Bromomethane			0.327	0.268	0.266	0.256	0.250	0.273	11.20
7) Chloroethane			0.286	0.248	0.223	0.227	0.226	0.242	10.92
8) T Dichlorotetra...			1.306	1.283	1.274	1.214	1.211	1.257	3.40
9) T Propene			0.734	0.736	0.698	0.660	0.630	0.691	6.69
10) T Heptane			1.847	1.940	1.897	1.792	1.719	1.839	4.73
11) T Trichlorofluor...			1.562	1.457	1.480	1.339	1.346	1.437	6.59
12) T 1,1,2-Trichlor...			1.247	1.215	1.188	1.120	1.135	1.181	4.52
13) Ethanol			0.110	0.086	0.073	0.029	0.028	0.065	55.48
14) T Bromoethene			0.529	0.447	0.460	0.417	0.409	0.452	10.47
15) T Acetone			1.699	1.573	1.469	1.113	1.139	1.399	18.73
16) T 1,3-Butadiene			0.952	0.705	0.608	0.573	0.577	0.683	23.38
17) tert-Butyl alc...			1.935	1.799	1.756	1.618	1.601	1.742	7.91
18) T 1,1-Dichloroet...			0.624	0.565	0.560	0.504	0.507	0.552	8.99
19) T Isopropyl Alcohol			1.016	0.800	0.738	0.702	0.689	0.789	16.99
20) T Methylene Chlo...			0.624	0.563	0.545	0.438	0.435	0.521	15.85
21) T Allyl Chloride			1.067	1.022	1.020	0.902	0.921	0.986	7.25
22) T trans-1,2-Dich...			0.704	0.701	0.621	0.590	0.575	0.638	9.56
23) T Vinyl Acetate			0.754	0.692	0.550	0.498	0.532	0.605	18.40
24) T 1,1-Dichloroet...			1.247	1.208	1.187	1.122	1.149	1.183	4.17
25) T Ethyl Acetate			4.017	3.619	3.703	3.457	3.353	3.630	7.06
26) T Hexane			1.507	1.462	1.383	1.333	1.325	1.402	5.71
27) T Carbon Disulfide			1.608	1.512	1.538	1.436	1.423	1.503	5.07
28) T Methyl tert-Bu...			0.921	0.886	0.894	0.831	0.832	0.873	4.56
29) T Chloroform			2.233	2.068	2.071	1.932	1.956	2.052	5.81
30) T Cyclohexane			1.337	1.238	1.211	1.162	1.140	1.217	6.33
31) T cis-1,2-Dichlo...			1.507	1.460	1.426	1.371	1.353	1.423	4.44
32) T 1,1,1-Trichlor...	3.151	2.495	2.279	2.201	2.152	2.062	2.105	2.349	16.23

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.754	0.771	0.770	0.667	0.664	0.725	7.58
35) T Carbon Tetrach...	1.069	0.782	0.709	0.732	0.748	0.703	0.725	0.781	16.60
36) T Benzene			1.049	1.027	1.013	0.934	0.934	0.991	5.47
37) T 1,2-Dichloroet...			0.613	0.569	0.563	0.537	0.556	0.568	4.95
38) T Trichloroethene	0.463	0.437	0.438	0.408	0.398	0.374	0.378	0.414	8.00
39) T 1,2-Dichloropr...			0.386	0.373	0.365	0.337	0.344	0.361	5.68

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\											
Method File : VL032725AIR.M											
40)	T	1,4-Dioxane			0.234	0.197	0.191	0.169	0.167	0.192	14.08
41)	T	Tetrahydrofuran			0.439	0.425	0.432	0.398	0.393	0.418	4.98
42)	T	Bromodichlorom...			0.778	0.779	0.802	0.769	0.785	0.782	1.58
43)		Methyl Methacr...			0.442	0.451	0.408	0.403	0.394	0.420	6.03
44)	T	2,2,4-Trimethy...			1.781	1.689	1.723	1.608	1.576	1.675	5.00
45)	T	t-1,3-Dichloro...			0.474	0.447	0.453	0.439	0.440	0.450	3.15
46)	T	cis-1,3-Dichlo...			0.549	0.560	0.566	0.550	0.535	0.552	2.12
47)	T	1,1,2-Trichlor...			0.419	0.382	0.379	0.364	0.365	0.382	5.88
48)	T	Dibromochlorom...			0.651	0.641	0.663	0.636	0.644	0.647	1.62
49)	T	Bromoform			0.566	0.588	0.597	0.582	0.570	0.581	2.18
50)	T	4-Methyl-2-Pen...			1.286	1.214	1.211	1.027	1.008	1.149	10.79
51)	T	2-Hexanone			0.999	1.018	1.031	0.844	0.827	0.944	10.56
52)	T	Tetrachloroethene	0.572	0.359	0.401	0.361	0.371	0.343	0.345	0.393	20.64
53)	T	Toluene			1.239	1.203	1.202	1.125	1.108	1.175	4.76
54)	T	1,2-Dibromoethane	0.574		0.543	0.538	0.520	0.508	0.510	0.532	4.74
-----ISTD-----											
55)	I	Chlorobenzene-d5									
56)		1,1,1,2-Tetrac...			0.576	0.575	0.576	0.510	0.515	0.550	6.30
57)	T	Chlorobenzene			1.024	0.979	0.994	0.875	0.871	0.949	7.46
58)	T	Ethyl Benzene			1.850	1.822	1.894	1.671	1.645	1.776	6.28
59)	T	m/p-Xylene			1.505	1.472	1.484	1.321	1.309	1.418	6.71
60)	T	o-Xylene			1.521	1.475	1.499	1.328	1.287	1.422	7.50
61)	T	Styrene			0.709	0.722	0.735	0.683	0.668	0.703	3.92
62)		Isopropylbenzene			2.290	2.199	2.261	1.962	1.920	2.126	8.14
63)	T	1,1,2,2-Tetrac...	0.986	0.800	0.907	0.889	0.893	0.796	0.777	0.864	8.78
64)		n-propylbenzene			0.578	0.581	0.576	0.499	0.502	0.547	7.81
65)		tert-Butylbenzene			2.190	2.071	2.144	1.736	1.691	1.966	11.96
66)	T	Benzyl Chloride			0.217	0.213	0.221	0.203	0.189	0.209	6.06
67)		sec-Butylbenzene			3.005	2.994	2.947	2.404	2.316	2.733	12.54
68)	S	1-Bromo-4-Fluo...	0.820	0.827	0.832	0.810	0.828	0.813	0.820	0.821	0.98
69)		p-Isopropyltol...			2.554	2.551	2.560	2.078	2.025	2.354	11.74
70)		n-Butylbenzene			2.536	2.598	2.682	2.137	2.082	2.407	11.51
71)		2-Chlorotoluene			1.717	1.728	1.759	1.538	1.510	1.650	7.07
72)	T	4-Ethyltoluene			1.834	1.859	1.929	1.629	1.601	1.770	8.26
73)	T	1,3,5-Trimethy...			1.616	1.671	1.677	1.387	1.381	1.546	9.71
74)	T	1,2,4-Trimethy...			1.994	1.958	1.966	1.510	1.478	1.781	14.74
75)	T	1,3-Dichlorobe...			1.036	1.022	1.063	0.882	0.882	0.977	8.99
76)	T	1,4-Dichlorobe...			1.030	1.017	1.040	0.888	0.886	0.972	8.05
77)	T	1,2-Dichlorobe...			1.012	1.031	1.014	0.850	0.850	0.951	9.78
78)	T	Hexachloro-1,3...			1.057	1.113	1.086	0.802	0.784	0.968	16.65
79)	T	Naphthalene	1.637		1.979	2.181	2.314	1.806	1.773	1.948	13.31
80)	T	Naphthalene,2-...			0.721	0.891	1.163	0.990	1.013	0.955	17.09
81)	T	1,2,4-Trichlor...			0.821	0.941	1.028	0.859	0.843	0.899	9.51

(#) = Out of Range

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1691 SAS No.: Q1691 SDG No.: Q1691

Instrument ID: MSVOA_L Calibration Date/Time: 04/04/2025 08:58

Lab File ID: VL042215.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	1.476	1.405		-4.81	30
Chloromethane	0.603	0.556		-7.79	30
Vinyl Chloride	0.639	0.525		-17.84	30
Bromomethane	0.273	0.237		-13.19	30
Chloroethane	0.242	0.212		-12.4	30
Tetrahydrofuran	0.418	0.366		-12.44	30
Trichlorofluoromethane	1.437	1.368		-4.8	30
1,1,2-Trichlorotrifluoroethane	1.181	1.042		-11.77	30
Dichlorotetrafluoroethane	1.257	1.215		-3.34	30
tert-Butyl alcohol	1.742	1.412		-18.94	30
Heptane	1.839	1.552		-15.61	30
1,1-Dichloroethene	0.552	0.469		-15.04	30
Acetone	1.399	1.172		-16.23	30
Carbon Disulfide	1.503	1.304		-13.24	30
Methyl tert-Butyl Ether	0.873	0.700		-19.82	30
Methylene Chloride	0.521	0.407		-21.88	30
trans-1,2-Dichloroethene	0.638	0.522		-18.18	30
1,1-Dichloroethane	1.183	1.062		-10.23	30
Cyclohexane	1.217	1.048		-13.89	30
2-Butanone	0.725	0.669		-7.72	30
Carbon Tetrachloride	0.781	0.794		1.66	30
cis-1,2-Dichloroethene	1.423	1.299		-8.71	30
Chloroform	2.052	2.021		-1.51	30
1,1,1-Trichloroethane	2.349	2.143		-8.77	30
2,2,4-Trimethylpentane	1.675	1.519		-9.31	30
Benzene	0.991	0.900		-9.18	30
1,2-Dichloroethane	0.568	0.591		4.05	30
Trichloroethene	0.414	0.384		-7.25	30
1,2-Dichloropropane	0.361	0.328		-9.14	30
Bromodichloromethane	0.782	0.802		2.56	30
4-Methyl-2-Pentanone	1.149	0.935		-18.63	30
Toluene	1.175	1.056		-10.13	30
t-1,3-Dichloropropene	0.450	0.379		-15.78	30
cis-1,3-Dichloropropene	0.552	0.479		-13.23	30
1,1,2-Trichloroethane	0.382	0.364		-4.71	30
Dibromochloromethane	0.647	0.652		0.77	30
1,2-Dibromoethane	0.532	0.505		-5.07	30
Tetrachloroethene	0.393	0.328		-16.54	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1691 SAS No.: Q1691 SDG No.: Q1691

Instrument ID: MSVOA_L Calibration Date/Time: 04/04/2025 08:58

Lab File ID: VL042215.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chlorobenzene	0.949	0.937		-1.26	30
Ethyl Benzene	1.776	1.694		-4.62	30
m/p-Xylene	1.418	1.420		0.14	30
o-Xylene	1.422	1.395		-1.9	30
Styrene	0.703	0.661		-5.97	30
Bromoform	0.581	0.562		-3.27	30
1,1,2,2-Tetrachloroethane	0.864	0.871		0.81	30
2-Chlorotoluene	1.650	1.655		0.3	30
1,3,5-Trimethylbenzene	1.546	1.558		0.78	30
1,2,4-Trimethylbenzene	1.781	1.730		-2.86	30
1,3-Dichlorobenzene	0.977	1.009		3.28	30
1,4-Dichlorobenzene	0.972	1.006		3.5	30
1,2-Dichlorobenzene	0.951	0.990		4.1	30
1,2,4-Trichlorobenzene	0.899	0.896		-0.33	30
Hexachloro-1,3-Butadiene	0.968	0.830		-14.26	30
1,3-Butadiene	0.683	0.570		-16.55	30
Naphthalene	1.948	2.004		2.88	30
4-Ethyltoluene	1.770	1.797		1.52	30
1-Bromo-4-Fluorobenzene	0.821	0.826		0.61	30
Hexane	1.402	1.237		-11.77	30
Allyl Chloride	0.986	0.844		-14.4	30
1,4-Dioxane	191.626	156.368		-18.4	30
Methyl Methacrylate	0.420	0.356		-15.24	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



SHIPPING DOCUMENTS

Client Contact Information				Bottle Order ID : B2503034				Courier : FGALDUN				3 of 6 COCs				
Client ID : GFELO1 Project ID : 10, Shore St.				Sampler Name(s) : FRANK GALDUN				Analysis				Matrix				
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : FRANK GALDUN				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified								
				Phone Number : 646-542-3465												
				Fax Number : 973-334-1692												
				Site Details: 416 CLINTON, ST. HEMPSTEAD, NY												
Analysis Turnaround Time				Standard : 10 business days OR				Data Package Type : LEVEL 2								
Rush (Specify): Days				EDD Type : RDE												
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
IAZR	3/21/25	7:47	3:47	30	4	65	68	-30	-4.7	10231	10024	6 L	12.5	VL042147.D	1	1
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) [Signature]						
		Ambient	Maximum	Minimum												
Start																
Stop																
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.						
		Ambient	Maximum	Minimum												
Start																
Stop																
Special Instructions/QC Requirements & Comments :																
Suspected Contamination: High Medium Low PID Readings 0.0																
Sampling site (State):																
Quick Connector required : NO																
Canisters Shipped by: Sam				Date/Time: 3/28/25				Canisters Received by: CD				Date/Time: 4-1-25 1010				
Samples Relinquished by: [Signature]				Date/Time:				Received by:				Date/Time:				
Relinquished by:				Date/Time:				Received by:				Date/Time:				

B2503034 - 1

Client Contact Information				Bottle Order ID : B2503034				Courier : FGALDUN				4 of 6 COCs					
Client ID : GFELO1				Project ID : 10 Eldridge St.				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix			
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : FRANK GALDUN				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 646-542-3465													
				Fax Number : 973-334-1692													
				Site Details: 416 CLINTON ST. HEMPSTEAD, NY													
				Analysis Turnaround Time				Data Package Type : LEVEL 2									
				Standard : 10 business days OR				EDD Type : RF									
Rush (Specify): Days																	
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor Ambient Air	Soil Gas
IAB	2/3/25	7:44	3:44	OVER 30	6	60	70	-30	-8-4	10581	10158	6 L	12.5	VL042147.D	1	1	
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) Syud							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings: 0,0																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by: SEM				Date/Time: 03/28/25				Canisters Received by: AO				Date/Time: 4-1-25 10:10				B2503034 - 2	
Samples Relinquished by: FRANK				Date/Time:				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by:				Date/Time:					

Client Contact Information				Bottle Order ID : B2503034				Courier : <u>FGALDUN</u>				<u>5</u> of <u>6</u> COCs																	
Client ID : GFEL01 Project ID : <u>19 Eldridge St.</u>				Sampler Name(s) : <u>FRANK GALDUN</u>				Analysis				Matrix																	
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : FRANK GALDUN				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																					
				Phone Number : 646-542-3465																									
				Fax Number : 973-334-1692																									
				Site Details: <u>416 CLINTON ST. HEMPSTEAD, NY</u>																									
Analysis Turnaround Time				Standard : <u>10 business days</u> OR				Data Package Type : <u>LEVEL 2</u>																					
Rush (Specify): _____ Days				EDD Type : <u>PDF</u>																									
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	<u>TO-15</u>	<u>Indoor Ambient Air</u>	<u>Soil Gas</u>												
<u>IA3DUP</u>	<u>3/31/25</u>	<u>7:43</u>	<u>3:43</u>	<u>30</u>	<u>5</u>	<u>68</u>	<u>70</u>	<u>-30</u>	<u>-7.1</u>	<u>10690</u>	<u>10326</u>	<u>6 L</u>	<u>12.5</u>	<u>VL042147.D</u>	<u>1</u>	<u>1</u>													
Temperature (Fahrenheit) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				GC/MS Analyst Signature (TO-15) <u>[Signature]</u>							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Pressure (Inches of Hg) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Special Instructions/QC Requirements & Comments :																													
Suspected Contamination: High Medium <u>Low</u> PID Readings: <u>0.0</u>																													
Sampling site (State):																													
Quick Connector required : <u>No</u>																													
Canisters Shipped by: <u>[Signature]</u>				Date/Time: <u>03/28/25</u>				Canisters Received by: <u>CF</u>				Date/Time: <u>4-1-25 1010</u>																	
Samples Relinquished by: <u>[Signature]</u>				Date/Time:				Received by:				Date/Time:																	
Relinquished by:				Date/Time:				Received by:				Date/Time:																	
B2503034 - 5																													

Client Contact Information				Bottle Order ID : B2503034				Courier : FGALDUN				6 of 6 COCs					
Client ID : GFELO1				Project ID : 10 Clinton St.				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix			
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : FRANK GALDUN				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 646-542-3465													
				Fax Number : 973-334-1692													
				Site Details: 416 CLINTON ST. HEMPSTEAD													
				Analysis Turnaround Time				Data Package Type : LEVEL 2									
				Standard : 10 business days OR				EDD Type : PDF									
Country :				Rush (Specify): Days													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
OAI	3/31/25	7:41	3:41	0.5	5.5			-30	-39	10227	10197	6 L	12.5	VL042147.D	1	1	
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 							
		Ambient	Maximum	Minimum													
Start		48															
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings: 0.0																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by: Sam				Date/Time: 3/28/25				Canisters Received by: CR				Date/Time: 4-1-25 1010				B2503034 - 3	
Samples Relinquished by: FRANK				Date/Time:				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by:				Date/Time:					

Laboratory Certification


Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

Internal Chain of Custody

Instructions: Use 1 form for each 20 samples of aliquot

Laboratory Person Breaking Field Seal on Sample Shuttle & Accepting Responsibility for Sample			
Laboratory: <u>Chemtech</u>		Location: <u>284 Sheffield Street, Mountainside, NJ 7092</u>	
NORGE		Title: <u>Sample Custodian</u>	
Field Sample Seal No.: <u>Q1691</u>		Date Broken: <u>4/1/2025</u>	Military Time Seal Broken: <u>10:10:00</u>
Case No.: <u>416 Clinton St Hempstead</u>		Analytical Parameter/Fraction: <u>TO-15</u>	

Sample No.	Aliquot/Extract No.	Sample No.	Aliquot/Extract No.
Q1691-01	IA2R		
Q1691-02	IA3		
Q1691-03	IA3DUP		
Q1691-04	OA1		

Date	Time	Relinquished By	Received By	Purpose of Change of Custody
4/1	1335	Signature 	Signature 	
		Printed Name <u>George N.</u>	Printed Name <u>George N.</u>	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	

Distribution: White - Original (Sent With Report)

Yellow - Contractor Archive

Pink - Sample Custodian - Interim Copy