

Duplicate Sample Summary

Lab Sample Id :	Q1691-01DUP	Q1691-01
Client Id :	IA2RDUP	IA2R
DF :	1	1
Datafile :	VL042222.D	VL042221.D
Anal Date & Time :	04/04/2025 13:20	04/04/2025 12:46

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0.19	0.25	27.3 *
2,2,4-Trimethylpentane	0	0	0
2-Butanone	1.1	1.1	0
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	9.7	9.5	2.1
Allyl Chloride	0	0	0
Benzene	0.18	0.18	0
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.11	0.12	8.7

Duplicate Sample Summary

Lab Sample Id :	Q1691-01DUP	Q1691-01
Client Id :	IA2RDUP	IA2R
DF :	1	1
Datafile :	VL042222.D	VL042221.D
Anal Date & Time :	04/04/2025 13:20	04/04/2025 12:46

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0.84	0.82	2.4
Chloromethane	0.96	0.93	3.2
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.41	0.41	0
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0	0	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0.42	0.45	6.9
m/p-Xylene	0.17	0	200 *
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	1	1.2	18.2
Naphthalene	0	0	0
o-Xylene	0	0	0
Styrene	0	0	0
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0.27	0.22	20.4 *
Tetrahydrofuran	0.1	0	200 *
Toluene	0.28	0.28	0
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.23	0.21	9.1
Vinyl Chloride	0	0	0