## ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

MATRIX: Water

CHEMTECH PROJECT NUMBER: Q1697

METHOD: 8270-Modified/3510 NA NO YES 1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) 2. GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ) 3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series. GC/MS Calibration - Initial Calibration performed within 30 days before sample 4. analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series. 5. GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration File ID BN036817.D met the requirements except for Fluoranthene-d10, The failure compound not associated with the client parameters list, therefore no corrective action was taken. 6. Blank Contamination - If yes, list compounds and concentrations in each blank: 7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries met the acceptable criteria except for IW-03-55-040125 [Terphenyl-d14 - 133%], this compound did not meet the NJDKQP criteria but met the in-house criteria. 8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples. The Blank Spike Duplicate met requirements for all samples . 9. Internal Standard Area/Retention Time Shift Meet Criteria Comments:

## ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA

NO

YES

10.	Extraction Holding Time Met		<b>✓</b>
	If not met, list number of days exceeded for each sample:		
11.	Analysis Holding Time Met		<b>✓</b>
	If not met, list number of days exceeded for each sample:		
ADDITI	IONAL COMMENTS:		
Sample	MW-19B-72-040125 was diluted due to high concentration.		
The For	m 6 is not included in the data package because the Initial Calibration was	s performed using 7 points.	
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %			
RSD val	tue for a compound is $<$ 20% for the Initial Calibration curve and use %D	calculated based on Amount added	
and Calo	culated amount for all compounds using Linear Regression when the %RS	SD value for a compound is > 20%	
for the I	nitial Calibration curve for SW-846 analysis.		
0 + DET			
QA REV	/IEW	Date	