

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc.	SDG No.: Q1697
Project: Former Schlumberger STC PTC Site # D3868221	Sample ID: Q1697-01
Client ID: MW-19B-72-040125DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Alkalinity	mg/L	+/-20	82.8		83.5		1	1		04/02/2025

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc. Project: Former Schlumberger STC PTC Site # D3868221 Client ID: IW-03-55-040125DUP	SDG No.: Q1697 Sample ID: Q1697-05 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	87.0		89.0		1	2.27		04/02/2025

Duplicate Sample Summary

Client: JACOBS Engineering Group, Inc.	SDG No.: Q1697
Project: Former Schlumberger STC PTC Site # D3868221	Sample ID: Q1697-05
Client ID: IW-03-55-040125MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-15	2.30		2.30		1	0		04/02/2025
Nitrate	mg/L	+/-15	2.50		2.50		1	0		04/02/2025
Nitrite	mg/L	+/-15	3.00		3.00		1	0		04/02/2025
Bromide	mg/L	+/-15	10.1		10.0		1	1		04/02/2025
Sulfate	mg/L	+/-15	17.9		17.7		1	1		04/02/2025
Chloride	mg/L	+/-15	16.4	OR	16.3	OR	1	1		04/02/2025
Orthophosphate as P	mg/L	+/-15	3.80		3.70		1	3		04/02/2025