



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1697

Contract: JACO05

Lab Code: CHEM

Case No.: Q1697

SAS No.: Q1697

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Aluminum	18.1	20.0	91	70 - 130	P	04/03/2025	16:41	LB135293
	Antimony	2.08	2.0	104	70 - 130	P	04/03/2025	16:41	LB135293
	Arsenic	1.00	1.0	100	70 - 130	P	04/03/2025	16:41	LB135293
	Barium	10.2	10.0	102	70 - 130	P	04/03/2025	16:41	LB135293
	Beryllium	1.07	1.0	107	70 - 130	P	04/03/2025	16:41	LB135293
	Cadmium	1.11	1.0	111	70 - 130	P	04/03/2025	16:41	LB135293
	Chromium	1.89	2.0	94	70 - 130	P	04/03/2025	16:41	LB135293
	Copper	1.86	2.0	93	70 - 130	P	04/03/2025	16:41	LB135293
	Iron	50.5	50.0	101	70 - 130	P	04/03/2025	16:41	LB135293
	Lead	1.00	1.0	100	70 - 130	P	04/03/2025	16:41	LB135293
	Magnesium	523	500	104	70 - 130	P	04/03/2025	16:41	LB135293
	Manganese	0.92	1.0	92	50 - 150	P	04/03/2025	16:41	LB135293
	Potassium	453	500	91	70 - 130	P	04/03/2025	16:41	LB135293
	Selenium	5.00	5.0	100	70 - 130	P	04/03/2025	16:41	LB135293
	Sodium	559	500	112	70 - 130	P	04/03/2025	16:41	LB135293