



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1697

Contract: JACO05 Lab Code: CHEM

Case No.: Q1697

SAS No.: Q1697

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	20.0	+/-20.0	U	20.0	P	04/03/2025	15:55	LB135293
	Antimony	2.00	+/-2.00	U	2.00	P	04/03/2025	15:55	LB135293
	Arsenic	1.00	+/-1.00	U	1.00	P	04/03/2025	15:55	LB135293
	Barium	10.0	+/-10.0	U	10.0	P	04/03/2025	15:55	LB135293
	Beryllium	1.00	+/-1.00	U	1.00	P	04/03/2025	15:55	LB135293
	Cadmium	1.00	+/-1.00	U	1.00	P	04/03/2025	15:55	LB135293
	Chromium	2.00	+/-2.00	U	2.00	P	04/03/2025	15:55	LB135293
	Copper	2.00	+/-2.00	U	2.00	P	04/03/2025	15:55	LB135293
	Iron	50.0	+/-50.0	U	50.0	P	04/03/2025	15:55	LB135293
	Lead	1.00	+/-1.00	U	1.00	P	04/03/2025	15:55	LB135293
	Magnesium	500	+/-500	U	500	P	04/03/2025	15:55	LB135293
	Manganese	1.00	+/-1.00	U	1.00	P	04/03/2025	15:55	LB135293
	Potassium	500	+/-500	U	500	P	04/03/2025	15:55	LB135293
	Selenium	5.00	+/-5.00	U	5.00	P	04/03/2025	15:55	LB135293
	Sodium	500	+/-500	U	500	P	04/03/2025	15:55	LB135293

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1697

Contract: JACO05

Lab Code: CHEM

Case No.: Q1697

SAS No.: Q1697

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	3.94	+/-20.0	J	20.0	P	04/03/2025	16:38	LB135293
	Antimony	2.00	+/-2.00	U	2.00	P	04/03/2025	16:38	LB135293
	Arsenic	1.00	+/-1.00	U	1.00	P	04/03/2025	16:38	LB135293
	Barium	10.0	+/-10.0	U	10.0	P	04/03/2025	16:38	LB135293
	Beryllium	1.00	+/-1.00	U	1.00	P	04/03/2025	16:38	LB135293
	Cadmium	1.00	+/-1.00	U	1.00	P	04/03/2025	16:38	LB135293
	Chromium	2.00	+/-2.00	U	2.00	P	04/03/2025	16:38	LB135293
	Copper	2.00	+/-2.00	U	2.00	P	04/03/2025	16:38	LB135293
	Iron	50.0	+/-50.0	U	50.0	P	04/03/2025	16:38	LB135293
	Lead	1.00	+/-1.00	U	1.00	P	04/03/2025	16:38	LB135293
	Magnesium	500	+/-500	U	500	P	04/03/2025	16:38	LB135293
	Manganese	1.00	+/-1.00	U	1.00	P	04/03/2025	16:38	LB135293
	Potassium	44.4	+/-500	J	500	P	04/03/2025	16:38	LB135293
	Selenium	5.00	+/-5.00	U	5.00	P	04/03/2025	16:38	LB135293
	Sodium	500	+/-500	U	500	P	04/03/2025	16:38	LB135293
CCB02	Aluminum	3.66	+/-20.0	J	20.0	P	04/03/2025	17:20	LB135293
	Antimony	2.00	+/-2.00	U	2.00	P	04/03/2025	17:20	LB135293
	Arsenic	1.00	+/-1.00	U	1.00	P	04/03/2025	17:20	LB135293
	Barium	10.0	+/-10.0	U	10.0	P	04/03/2025	17:20	LB135293
	Beryllium	1.00	+/-1.00	U	1.00	P	04/03/2025	17:20	LB135293
	Cadmium	1.00	+/-1.00	U	1.00	P	04/03/2025	17:20	LB135293
	Chromium	2.00	+/-2.00	U	2.00	P	04/03/2025	17:20	LB135293
	Copper	2.00	+/-2.00	U	2.00	P	04/03/2025	17:20	LB135293
	Iron	50.0	+/-50.0	U	50.0	P	04/03/2025	17:20	LB135293
	Lead	1.00	+/-1.00	U	1.00	P	04/03/2025	17:20	LB135293
	Magnesium	500	+/-500	U	500	P	04/03/2025	17:20	LB135293
	Manganese	1.00	+/-1.00	U	1.00	P	04/03/2025	17:20	LB135293
	Potassium	500	+/-500	U	500	P	04/03/2025	17:20	LB135293
	Selenium	5.00	+/-5.00	U	5.00	P	04/03/2025	17:20	LB135293
	Sodium	500	+/-500	U	500	P	04/03/2025	17:20	LB135293
CCB03	Aluminum	20.0	+/-20.0	U	20.0	P	04/03/2025	18:41	LB135293
	Antimony	2.00	+/-2.00	U	2.00	P	04/03/2025	18:41	LB135293
	Arsenic	1.00	+/-1.00	U	1.00	P	04/03/2025	18:41	LB135293
	Barium	10.0	+/-10.0	U	10.0	P	04/03/2025	18:41	LB135293
	Beryllium	1.00	+/-1.00	U	1.00	P	04/03/2025	18:41	LB135293
	Cadmium	1.00	+/-1.00	U	1.00	P	04/03/2025	18:41	LB135293
	Chromium	2.00	+/-2.00	U	2.00	P	04/03/2025	18:41	LB135293
	Copper	2.00	+/-2.00	U	2.00	P	04/03/2025	18:41	LB135293
	Iron	50.0	+/-50.0	U	50.0	P	04/03/2025	18:41	LB135293

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1697

Contract: JACO05

Lab Code: CHEM

Case No.: Q1697

SAS No.: Q1697

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Lead	1.00	+/-1.00	U	1.00	P	04/03/2025	18:41	LB135293
	Magnesium	500	+/-500	U	500	P	04/03/2025	18:41	LB135293
	Manganese	0.93	+/-1.00	J	1.00	P	04/03/2025	18:41	LB135293
	Potassium	500	+/-500	U	500	P	04/03/2025	18:41	LB135293
	Selenium	5.00	+/-5.00	U	5.00	P	04/03/2025	18:41	LB135293
	Sodium	500	+/-500	U	500	P	04/03/2025	18:41	LB135293