

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1697
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1697 **SAS No.:** Q1697
Matrix: Water **Level:** LOW **Client ID:** IW-02-55-040125-FDA
Sample ID: Q1697-04 **Spiked ID:** Q1697-04A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	0.53	J	0.39	J	500	0		P
Potassium	ug/L	75 - 125	25400		7150		25000	73		P