

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1697
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1697 **SAS No.:** Q1697
Matrix: Water **Sample ID:** Q1697-04 **Client ID:** IW-02-55-040125-FDDUP
Percent Solids for Sample: NA **Duplicate ID** Q1697-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	70.8		58.6		19		P
Antimony	ug/L	20	2.00	U	2.00	U			P
Arsenic	ug/L	20	0.39	J	0.40	J	3		P
Barium	ug/L	20	303		303		0		P
Beryllium	ug/L	20	1.00	U	1.00	U			P
Cadmium	ug/L	20	1.00	U	1.00	U			P
Chromium	ug/L	20	0.86	J	0.59	J	37		P
Copper	ug/L	20	2.00	U	2.00	U			P
Iron	ug/L	20	5440		5400		1		P
Lead	ug/L	20	1.00	U	1.00	U			P
Magnesium	ug/L	20	8590		8620		0		P
Manganese	ug/L	20	602		603		0		P
Potassium	ug/L	20	7150		7030		2		P
Selenium	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	6930		6980		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** Q1697
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1697 **SAS No.:** Q1697
Matrix: Water **Sample ID:** Q1697-04MS **Client ID:** IW-02-55-040125-FDMSD
Percent Solids for Sample: NA **Duplicate ID** Q1697-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	9340		9350		0		P
Antimony	ug/L	20	485		490		1		P
Arsenic	ug/L	20	0.54	J	0.56	J	4		P
Barium	ug/L	20	2410		2450		2		P
Beryllium	ug/L	20	480		484		1		P
Cadmium	ug/L	20	481		487		1		P
Chromium	ug/L	20	468		472		1		P
Copper	ug/L	20	4300		4360		1		P
Iron	ug/L	20	27300		27400		0		P
Lead	ug/L	20	2380		2360		1		P
Magnesium	ug/L	20	58200		58900		1		P
Manganese	ug/L	20	4900		4950		1		P
Potassium	ug/L	20	24800		25200		2		P
Selenium	ug/L	20	472		471		0		P
Sodium	ug/L	20	57500		57800		1		P