

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1697

Contract: JACO05

Lab Code: CHEM

Case No.: Q1697

SAS No.: Q1697

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167437BS							
Aluminum	ug/L	10000	10800		108	80 - 120	P
Antimony	ug/L	500	527		105	80 - 120	P
Arsenic	ug/L	500	518		104	80 - 120	P
Barium	ug/L	2500	2610		104	80 - 120	P
Beryllium	ug/L	500	519		104	80 - 120	P
Cadmium	ug/L	500	535		107	80 - 120	P
Chromium	ug/L	500	523		105	80 - 120	P
Copper	ug/L	5000	5240		105	80 - 120	P
Iron	ug/L	25000	27500		110	80 - 120	P
Lead	ug/L	2500	2540		102	80 - 120	P
Magnesium	ug/L	50000	55200		110	80 - 120	P
Manganese	ug/L	5000	5270		105	80 - 120	P
Potassium	ug/L	25000	22400		90	80 - 120	P
Selenium	ug/L	500	514		103	80 - 120	P
Sodium	ug/L	50000	56000		112	80 - 120	P