

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1697

Sas No.: Q1697 SDG No.: Q1697

Instrument ID: P8

Start Date : 04/03/2025

Run Number: LB135293

End Date : 04/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1415	100		100							
S2	S2	1422	99		100							
S3	S3	1425	98		99							
S4	S4	1428	99		98							
S5	S5	1431	101		98							
S6	S6	1434	99		95							
S7	S7	1437	98		93							
S8	S8	1440	85		75							
ICV01	ICV01	1526	96		96							
LLICV	LLICV	1550	96		98							
ICB01	ICB01	1555	96		97							
ICSA01	ICSA01	1603	93		86							
ICSAB01	ICSAB01	1620	97		89							
CCV01	CCV01	1635	92		83							
CCB01	CCB01	1638	100		101							
CRI	CRI	1641	102		100							
PB167437BL	PB167437BL	1645	102		101							
PB167437BS	PB167437BS	1648	99		93							
Q1697-01	MW-19B-72-04	1652	102		100							
Q1697-02	IW-01-55-04C	1656	102		100							
Q1697-03	IW-02-55-04C	1659	103		99							
Q1697-04	IW-02-55-04C	1702	102		100							
Q1697-04DUP	IW-02-55-04C	1705	100		99							
Q1697-04L	IW-02-55-04C	1708	100		99							
Q1697-04MS	IW-02-55-04C	1712	99		94							
CCV02	CCV02	1716	94		84							
CCB02	CCB02	1720	104		103							
Q1697-04MSD	IW-02-55-04C	1723	101		96							
Q1697-04A	IW-02-55-04C	1726	100		96							
Q1697-05	IW-03-55-04C	1729	104		103							
Q1697-13	MW-19B-72-04	1732	103		100							
Q1697-14	IW-01-55-04C	1735	103		100							
Q1697-15	IW-02-55-04C	1738	103		100							
Q1697-16	IW-02-55-04C	1742	101		98							

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1697

Sas No.: Q1697 SDG No.: Q1697

Instrument ID: P8

Start Date : 04/03/2025

Run Number: LB135293

End Date : 04/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
Q1697-17	IW-03-55-040	1745	102		98							
CCV03	CCV03	1752	90		82							
CCB03	CCB03	1841	108		108							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JAC005

Lab Code: CHEM Case no.: Q1697

Sas No.: Q1697 SDG No.: Q1697

Instrument ID: P8

Start Date : 04/03/2025

Run Number: LB135293

End Date : 04/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1415	100									
S2	S2	1422	100									
S3	S3	1425	98									
S4	S4	1428	96									
S5	S5	1431	94									
S6	S6	1434	91									
S7	S7	1437	90									
S8	S8	1440	77									
ICV01	ICV01	1526	97									
LLICV	LLICV	1550	99									
ICB01	ICB01	1555	98									
ICSA01	ICSA01	1603	86									
ICSAB01	ICSAB01	1620	86									
CCV01	CCV01	1635	84									
CCB01	CCB01	1638	100									
CRI	CRI	1641	101									
PB167437BL	PB167437BL	1645	100									
PB167437BS	PB167437BS	1648	92									
Q1697-01	MW-19B-72-04	1652	97									
Q1697-02	IW-01-55-04C	1656	98									
Q1697-03	IW-02-55-04C	1659	99									
Q1697-04	IW-02-55-04C	1702	99									
Q1697-04DUP	IW-02-55-04C	1705	98									
Q1697-04L	IW-02-55-04C	1708	99									
Q1697-04MS	IW-02-55-04C	1712	92									
CCV02	CCV02	1716	84									
CCB02	CCB02	1720	103									
Q1697-04MSD	IW-02-55-04C	1723	92									
Q1697-04A	IW-02-55-04C	1726	92									
Q1697-05	IW-03-55-04C	1729	101									
Q1697-13	MW-19B-72-04	1732	99									
Q1697-14	IW-01-55-04C	1735	100									
Q1697-15	IW-02-55-04C	1738	99									
Q1697-16	IW-02-55-04C	1742	97									

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: JACOBS Engineering Group, Inc.

Contract: JACO05

Lab Code: CHEM Case no.: Q1697

Sas No.: Q1697 SDG No.: Q1697

Instrument ID: P8

Start Date : 04/03/2025

Run Number: LB135293

End Date : 04/03/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
Q1697-17	IW-03-55-040	1745	99									
CCV03	CCV03	1752	85									
CCB03	CCB03	1841	100									